

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 9/30/2024 9:40:19 AM

## JOB DESCRIPTION

RED-HILL  
Weekly PFAS

## JOB NUMBER

380-113907-1

# Eurofins Eaton Analytical Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Generated  
9/30/2024 9:40:19 AM

Authorized for release by  
Rachelle Arada, Project Manager  
[Rachelle.Arada@et.eurofinsus.com](mailto:Rachelle.Arada@et.eurofinsus.com)  
(626)386-1106



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
Isotope Dilution Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	29
Lab Chronicle . . . . .	31
Certification Summary . . . . .	32
Method Summary . . . . .	33
Sample Summary . . . . .	34
Chain of Custody . . . . .	35
Receipt Checklists . . . . .	37

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Qualifiers

### LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-113907-1

**Job ID: 380-113907-1**

**Eurofins Eaton Analytical Pomona**

## Job Narrative 380-113907-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

### Receipt

The samples were received on 9/19/2024 9:59 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.1°C.

### PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Client Sample ID: HALAWA SHAFT VIEWING POOL

## Lab Sample ID: 380-113907-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.3		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.1		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.8		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.8		2.0	ng/L	1		537.1	Total/NA

## Client Sample ID: FB: HALAWA SHAFT VIEWING POOL

## Lab Sample ID: 380-113907-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-1**

Date Collected: 09/17/24 10:00

Matrix: Water

Date Received: 09/19/24 09:59

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.3</b>		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>3.1</b>		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:34	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	82		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C6 PFDA	69		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C5 PFHxA	89		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C4 PFHpA	81		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C8 PFOA	82		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C9 PFNA	76		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C7 PFUnA	78		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C2 PFDoA	81		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C4 PFBA	99		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C5 PFPeA	99		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C3 PFBS	98		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C3 PFHxS	100		50 - 200	09/27/24 10:53	09/29/24 11:34	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-1**

**Date Collected: 09/17/24 10:00**

**Matrix: Water**

**Date Received: 09/19/24 09:59**

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	105		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C2-4:2-FTS	96		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C2-6:2-FTS	94		50 - 200	09/27/24 10:53	09/29/24 11:34	1
13C2-8:2-FTS	89		50 - 200	09/27/24 10:53	09/29/24 11:34	1

**Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>3.8</b>		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.8</b>		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:45	1
Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac		
d5-NEtFOSAA	122		70 - 130	09/21/24 10:10	09/22/24 14:45	1		
13C2 PFHxA	118		70 - 130	09/21/24 10:10	09/22/24 14:45	1		
13C2 PFDA	117		70 - 130	09/21/24 10:10	09/22/24 14:45	1		
13C3-GenX	118		70 - 130	09/21/24 10:10	09/22/24 14:45	1		

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-2**

**Date Collected: 09/17/24 10:00**

**Matrix: Water**

**Date Received: 09/19/24 09:59**

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1

Eurofins Eaton Analytical Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-2**

Date Collected: 09/17/24 10:00

Matrix: Water

Date Received: 09/19/24 09:59

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		09/27/24 10:53	09/29/24 11:45	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	95		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C6 PFDA	101		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C5 PFHxA	111		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C4 PFHpA	105		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C8 PFOA	107		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C9 PFNA	103		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C7 PFUnA	108		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C2 PFDoA	107		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C4 PFBA	105		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C5 PFPeA	103		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C3 PFBS	102		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C3 PFHxS	104		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C8 PFOS	106		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C2-4:2-FTS	94		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C2-6:2-FTS	90		50 - 200	09/27/24 10:53	09/29/24 11:45	1
13C2-8:2-FTS	91		50 - 200	09/27/24 10:53	09/29/24 11:45	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-2**

**Date Collected: 09/17/24 10:00**

**Matrix: Water**

**Date Received: 09/19/24 09:59**

**Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		09/23/24 09:19	09/24/24 17:34	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	107		70 - 130			09/23/24 09:19	09/24/24 17:34	1
13C2 PFHxA	106		70 - 130			09/23/24 09:19	09/24/24 17:34	1
13C2 PFDA	101		70 - 130			09/23/24 09:19	09/24/24 17:34	1
13C3-GenX	95		70 - 130			09/23/24 09:19	09/24/24 17:34	1

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.3		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.1		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.8		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.8		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-2**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-113907-1  
 SDG: Weekly PFAS

**Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-113655-B-1-A MS	Matrix Spike	115	103	114	105
380-113655-C-1-A MSD	Matrix Spike Duplicate	111	100	107	105
380-113907-1	HALAWA SHAFT VIEWING POOL	122	118	117	118
380-113907-1 MS	HALAWA SHAFT VIEWING POOL	128	123	121	126
380-113907-1 MSD	HALAWA SHAFT VIEWING POOL	118	110	114	112
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	107	106	101	95
LCS 380-109544/24-A	Lab Control Sample	112	100	109	107
LCS 380-109690/22-A	Lab Control Sample	111	109	104	107
MBL 380-109544/22-A	Method Blank	130	111	111	117
MBL 380-109690/20-A	Method Blank	114	107	108	103
MRL 380-109544/23-A	Lab Control Sample	123	110	117	114
MRL 380-109690/21-A	Lab Control Sample	115	110	102	106

**Surrogate Legend**

- d5NEFOS = d5-NEtFOSAA
- PFHxA = 13C2 PFHxA
- PFDA = 13C2 PFDA
- GenX = 13C3-GenX



# Isotope Dilution Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-113854-B-1-A MS	Matrix Spike	83	67	90	81	77	69	73	78
380-113854-C-1-A MSD	Matrix Spike Duplicate	82	78	87	81	81	74	81	83
380-113907-1	HALAWA SHAFT VIEWING POOL	82	69	89	81	82	76	78	81
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	95	101	111	105	107	103	108	107
LCS 380-110608/21-A	Lab Control Sample	98	105	107	103	105	100	107	107
MBL 380-110608/19-A	Method Blank	85	98	97	93	97	96	102	98
MRL 380-110608/20-A	Lab Control Sample	87	102	102	97	104	101	105	103

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-113854-B-1-A MS	Matrix Spike	93	96	101	103	104	94	97	97
380-113854-C-1-A MSD	Matrix Spike Duplicate	90	98	100	97	101	89	97	95
380-113907-1	HALAWA SHAFT VIEWING POOL	99	99	98	100	105	96	94	89
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	105	103	102	104	106	94	90	91
LCS 380-110608/21-A	Lab Control Sample	106	108	95	102	103	89	95	94
MBL 380-110608/19-A	Method Blank	97	101	101	101	105	95	94	95
MRL 380-110608/20-A	Lab Control Sample	97	100	99	100	101	92	93	95

### Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDaA = 13C2 PFDaA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Lab Sample ID: MBL 380-110608/19-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.35		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1
Perfluoropentanesulfonic acid (PFPeS)	<0.38		2.0	ng/L		09/27/24 10:53	09/29/24 07:48	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	85		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C6 PFDA	98		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C5 PFHxA	97		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C4 PFHpA	93		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C8 PFOA	97		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C9 PFNA	96		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C7 PFUnA	102		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C2 PFDoA	98		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C4 PFBA	97		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C5 PFPeA	101		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C3 PFBS	101		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C3 PFHxS	101		50 - 200	09/27/24 10:53	09/29/24 07:48	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MBL 380-110608/19-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	105		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C2-4:2-FTS	95		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C2-6:2-FTS	94		50 - 200	09/27/24 10:53	09/29/24 07:48	1
13C2-8:2-FTS	95		50 - 200	09/27/24 10:53	09/29/24 07:48	1

**Lab Sample ID: LCS 380-110608/21-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	119		ng/L		100	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	116		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	119		ng/L		99	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	115		ng/L		96	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	118		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	120	111		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	120	119		ng/L		100	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	117		ng/L		97	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	111		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	120	112		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	120	119		ng/L		99	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	112		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	120	118		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	113		ng/L		94	70 - 130
Perfluorobutanoic acid (PFBA)	120	112		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	116		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	125		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	114		ng/L		95	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	104		ng/L		87	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	120	124		ng/L		103	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	109		ng/L		91	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	117		ng/L		98	70 - 130
Perfluoropentanoic acid (PFPeA)	120	114		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	116		ng/L		97	70 - 130

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: LCS 380-110608/21-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	113		ng/L		94	70 - 130
<b>LCS LCS</b>							
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				
13C3 HFPO-DA	98		50 - 200				
13C6 PFDA	105		50 - 200				
13C5 PFHxA	107		50 - 200				
13C4 PFHpA	103		50 - 200				
13C8 PFOA	105		50 - 200				
13C9 PFNA	100		50 - 200				
13C7 PFUnA	107		50 - 200				
13C2 PFDoA	107		50 - 200				
13C4 PFBA	106		50 - 200				
13C5 PFPeA	108		50 - 200				
13C3 PFBS	95		50 - 200				
13C3 PFHxS	102		50 - 200				
13C8 PFOS	103		50 - 200				
13C2-4:2-FTS	89		50 - 200				
13C2-6:2-FTS	95		50 - 200				
13C2-8:2-FTS	94		50 - 200				

**Lab Sample ID: MRL 380-110608/20-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.90	J	ng/L		95	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.00	J	ng/L		100	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.08	J	ng/L		104	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.93	J	ng/L		97	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.07	J	ng/L		104	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	1.91	J	ng/L		95	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.04	J	ng/L		102	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.11	J	ng/L		106	50 - 150

Eurofins Eaton Analytical Pomona



# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MRL 380-110608/20-A**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.22	J	ng/L		111	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.20	J	ng/L		110	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.37	J	ng/L		119	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	1.85	J	ng/L		93	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	1.85	J	ng/L		93	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	1.89	J	ng/L		95	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	1.92	J	ng/L		96	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.00	J	ng/L		100	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	1.91	J	ng/L		95	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	1.87	J	ng/L		94	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	87		50 - 200
13C6 PFDA	102		50 - 200
13C5 PFHxA	102		50 - 200
13C4 PFHpA	97		50 - 200
13C8 PFOA	104		50 - 200
13C9 PFNA	101		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	103		50 - 200
13C4 PFBA	97		50 - 200
13C5 PFPeA	100		50 - 200
13C3 PFBS	99		50 - 200
13C3 PFHxS	100		50 - 200
13C8 PFOS	101		50 - 200
13C2-4:2-FTS	92		50 - 200
13C2-6:2-FTS	93		50 - 200
13C2-8:2-FTS	95		50 - 200

**Lab Sample ID: 380-113854-B-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.4	59.6		ng/L		99	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.4	56.9		ng/L		94	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.4	60.0		ng/L		99	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-113854-B-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.4	57.6		ng/L		95	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	2.0		60.4	61.8		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.4	57.4		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.4	62.3		ng/L		103	70 - 130
Perfluoroheptanoic acid (PFHpA)	2.7		60.4	62.2		ng/L		99	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	2.8		60.4	59.6		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	3.5		60.4	60.2		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.4	59.2		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.4	56.3		ng/L		92	70 - 130
Perfluorooctanoic acid (PFOA)	5.8		60.4	64.1		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.4	61.5		ng/L		102	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.4	61.2		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.4	57.7		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.4	58.9		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.4	55.9		ng/L		93	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.4	52.0		ng/L		86	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.4	56.4		ng/L		93	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.4	57.7		ng/L		96	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.4	56.3		ng/L		93	70 - 130
Perfluoropentanoic acid (PFPeA)	3.5		60.4	58.4		ng/L		91	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.4	57.1		ng/L		95	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.4	56.7		ng/L		92	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	83		50 - 200
13C6 PFDA	67		50 - 200
13C5 PFHxA	90		50 - 200
13C4 PFHpA	81		50 - 200
13C8 PFOA	77		50 - 200
13C9 PFNA	69		50 - 200
13C7 PFUnA	73		50 - 200
13C2 PFDoA	78		50 - 200
13C4 PFBA	93		50 - 200
13C5 PFPeA	96		50 - 200
13C3 PFBS	101		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	104		50 - 200

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-113854-B-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	94		50 - 200
13C2-6:2-FTS	97		50 - 200
13C2-8:2-FTS	97		50 - 200

**Lab Sample ID: 380-113854-C-1-A MSD**  
**Matrix: Water**  
**Analysis Batch: 110795**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 110608**

<b>Analyte</b>	<b>Sample Result</b>	<b>Sample Qualifier</b>	<b>Spike Added</b>	<b>MSD Result</b>	<b>MSD Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec Limits</b>	<b>RPD</b>	<b>RPD Limit</b>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	59.8		ng/L		99	70 - 130	0	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	57.9		ng/L		96	70 - 130	2	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	59.0		ng/L		98	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.2	59.0		ng/L		98	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	2.0		60.2	60.9		ng/L		98	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		60.2	55.1		ng/L		91	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	61.8		ng/L		103	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	2.7		60.2	64.2		ng/L		102	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	2.8		60.2	62.0		ng/L		98	70 - 130	4	30
Perfluorohexanoic acid (PFHxA)	3.5		60.2	62.7		ng/L		98	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		60.2	61.6		ng/L		101	70 - 130	4	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.2	57.5		ng/L		95	70 - 130	2	30
Perfluorooctanoic acid (PFOA)	5.8		60.2	67.0		ng/L		102	70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	60.7		ng/L		101	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		60.2	64.4		ng/L		104	70 - 130	5	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	61.1		ng/L		101	70 - 130	6	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	61.4		ng/L		102	70 - 130	4	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	62.0		ng/L		103	70 - 130	10	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	60.2		ng/L		100	70 - 130	15	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		60.2	57.2		ng/L		95	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	59.0		ng/L		98	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	58.2		ng/L		97	70 - 130	3	30
Perfluoropentanoic acid (PFPeA)	3.5		60.2	60.2		ng/L		94	70 - 130	3	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	58.9		ng/L		98	70 - 130	3	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	60.3		ng/L		98	70 - 130	6	30

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	82		50 - 200
13C6 PFDA	78		50 - 200
13C5 PFHxA	87		50 - 200
13C4 PFHpA	81		50 - 200
13C8 PFOA	81		50 - 200
13C9 PFNA	74		50 - 200
13C7 PFUnA	81		50 - 200
13C2 PFDoA	83		50 - 200
13C4 PFBA	90		50 - 200
13C5 PFPeA	98		50 - 200
13C3 PFBS	100		50 - 200
13C3 PFHxS	97		50 - 200
13C8 PFOS	101		50 - 200
13C2-4:2-FTS	89		50 - 200
13C2-6:2-FTS	97		50 - 200
13C2-8:2-FTS	95		50 - 200

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-109544/22-A  
Matrix: Water  
Analysis Batch: 109606

Client Sample ID: Method Blank  
Prep Type: Total/NA  
Prep Batch: 109544

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.29		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		09/21/24 10:10	09/22/24 14:13	1
Surrogate	MBL MBL		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
d5-NEtFOSAA	130		70 - 130			09/21/24 10:10	09/22/24 14:13	1
13C2 PFHxA	111		70 - 130			09/21/24 10:10	09/22/24 14:13	1
13C2 PFDA	111		70 - 130			09/21/24 10:10	09/22/24 14:13	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: MBL 380-109544/22-A**  
**Matrix: Water**  
**Analysis Batch: 109606**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 109544**

Surrogate	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C3-GenX	117		70 - 130	09/21/24 10:10	09/22/24 14:13	1

**Lab Sample ID: LCS 380-109544/24-A**  
**Matrix: Water**  
**Analysis Batch: 109606**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 109544**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	25.0	25.6		ng/L		103	70 - 130
Perfluorooctanesulfonic acid (PFOS)	25.0	27.1		ng/L		109	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.0	27.0		ng/L		108	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.0	27.4		ng/L		109	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.0	29.6		ng/L		118	70 - 130
Perfluorohexanoic acid (PFHxA)	25.0	25.3		ng/L		101	70 - 130
Perfluorododecanoic acid (PFDoA)	25.0	27.6		ng/L		111	70 - 130
Perfluorooctanoic acid (PFOA)	25.0	26.8		ng/L		107	70 - 130
Perfluorodecanoic acid (PFDA)	25.0	26.4		ng/L		106	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.0	28.1		ng/L		112	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.0	23.7		ng/L		95	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.0	26.2		ng/L		105	70 - 130
Perfluorononanoic acid (PFNA)	25.0	26.3		ng/L		105	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.0	29.1		ng/L		116	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.0	28.7		ng/L		115	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	25.0	29.8		ng/L		119	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.0	27.1		ng/L		109	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.0	26.2		ng/L		105	70 - 130

  

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	112		70 - 130
13C2 PFHxA	100		70 - 130
13C2 PFDA	109		70 - 130
13C3-GenX	107		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: MRL 380-109544/23-A**  
**Matrix: Water**  
**Analysis Batch: 109606**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 109544**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.28	J	ng/L		114	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.30	J	ng/L		115	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.15	J	ng/L		107	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.34	J	ng/L		117	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.25	J	ng/L		112	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.31	J	ng/L		116	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.13	J	ng/L		106	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.26	J	ng/L		113	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.83	J	ng/L		92	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.24	J	ng/L		112	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	2.48	J	ng/L		124	50 - 150
Perfluorotridecanoic acid (PFTTrDA)	2.00	2.37	J	ng/L		118	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.24	J	ng/L		112	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.03	J	ng/L		101	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.23	J	ng/L		111	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	123		70 - 130
13C2 PFHxA	110		70 - 130
13C2 PFDA	117		70 - 130
13C3-GenX	114		70 - 130

**Lab Sample ID: 380-113907-1 MS**  
**Matrix: Water**  
**Analysis Batch: 109606**

**Client Sample ID: HALAWA SHAFT VIEWING POOL**  
**Prep Type: Total/NA**  
**Prep Batch: 109544**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.2	50.5		ng/L		101	70 - 130
Perfluorooctanesulfonic acid (PFOS)	3.8		50.2	55.9		ng/L		104	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.2	53.6		ng/L		107	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.2	54.4		ng/L		108	70 - 130

Eurofins Eaton Analytical Pomona







# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: MBL 380-109690/20-A**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		09/23/24 09:19	09/24/24 13:45	1

  

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	114		70 - 130	09/23/24 09:19	09/24/24 13:45	1
13C2 PFHxA	107		70 - 130	09/23/24 09:19	09/24/24 13:45	1
13C2 PFDA	108		70 - 130	09/23/24 09:19	09/24/24 13:45	1
13C3-GenX	103		70 - 130	09/23/24 09:19	09/24/24 13:45	1

**Lab Sample ID: LCS 380-109690/22-A**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	50.1	51.0		ng/L		102	70 - 130
Perfluorooctanesulfonic acid (PFOS)	50.1	52.8		ng/L		105	70 - 130
Perfluoroundecanoic acid (PFUnA)	50.1	47.2		ng/L		94	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.1	50.5		ng/L		101	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.1	51.5		ng/L		103	70 - 130
Perfluorohexanoic acid (PFHxA)	50.1	50.3		ng/L		100	70 - 130
Perfluorododecanoic acid (PFDoA)	50.1	47.5		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	50.1	51.3		ng/L		102	70 - 130
Perfluorodecanoic acid (PFDA)	50.1	48.7		ng/L		97	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.1	55.9		ng/L		112	70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.1	46.9		ng/L		94	70 - 130
Perfluoroheptanoic acid (PFHpA)	50.1	51.8		ng/L		103	70 - 130
Perfluorononanoic acid (PFNA)	50.1	48.3		ng/L		96	70 - 130
Perfluorotetradecanoic acid (PFTA)	50.1	47.1		ng/L		94	70 - 130
Perfluorotridecanoic acid (PFTTrDA)	50.1	46.7		ng/L		93	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	50.1	54.5		ng/L		109	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.1	52.5		ng/L		105	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.1	50.1		ng/L		100	70 - 130

  

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	111		70 - 130
13C2 PFHxA	109		70 - 130
13C2 PFDA	104		70 - 130

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: LCS 380-109690/22-A**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
13C3-GenX	107		70 - 130

**Lab Sample ID: MRL 380-109690/21-A**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Analyte	Spike Added	MRL MRL		Unit	D	%Rec	Limits
		Result	Qualifier				
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.07	J	ng/L		104	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	1.90	J	ng/L		95	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.13	J	ng/L		107	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.17	J	ng/L		109	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.24	J	ng/L		112	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.04	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.26	J	ng/L		113	50 - 150
Perfluorononanoic acid (PFNA)	2.00	1.97	J	ng/L		99	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	1.88	J	ng/L		94	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.00	2.12	J	ng/L		106	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.06	J	ng/L		103	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.05	J	ng/L		103	50 - 150

Surrogate	MRL MRL		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	115		70 - 130
13C2 PFHxA	110		70 - 130
13C2 PFDA	102		70 - 130
13C3-GenX	106		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: 380-113655-B-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec	Limits
	Result	Qualifier	Added	Result	Qualifier					
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.7	46.4		ng/L		91		70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.7	53.4		ng/L		103		70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.7	51.7		ng/L		102		70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.7	53.6		ng/L		106		70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		50.7	54.0		ng/L		107		70 - 130
Perfluorohexanoic acid (PFHxA)	13		50.7	61.0		ng/L		94		70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		50.7	49.5		ng/L		98		70 - 130
Perfluorooctanoic acid (PFOA)	4.9		50.7	59.1		ng/L		107		70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		50.7	52.7		ng/L		104		70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		50.7	55.0		ng/L		105		70 - 130
Perfluorobutanesulfonic acid (PFBS)	2.5		50.7	47.9		ng/L		95		70 - 130
Perfluoroheptanoic acid (PFHpA)	2.1		50.7	49.0		ng/L		92		70 - 130
Perfluorononanoic acid (PFNA)	<2.0		50.7	52.5		ng/L		103		70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		50.7	47.1		ng/L		93		70 - 130
Perfluorotridecanoic acid (PFTrDA)	<2.0		50.7	46.5		ng/L		92		70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		50.7	52.7		ng/L		104		70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		50.7	50.6		ng/L		100		70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		50.7	48.0		ng/L		95		70 - 130

Surrogate	MS MS		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	115		70 - 130
13C2 PFHxA	103		70 - 130
13C2 PFDA	114		70 - 130
13C3-GenX	105		70 - 130

**Lab Sample ID: 380-113655-C-1-A MSD**  
**Matrix: Water**  
**Analysis Batch: 109924**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 109690**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	Limits	RPD	
	Result	Qualifier	Added	Result	Qualifier						RPD	Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.2	46.7		ng/L		93		70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.2	55.8		ng/L		109		70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		50.2	49.5		ng/L		99		70 - 130	4	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.2	52.7		ng/L		105		70 - 130	2	30

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: 380-113655-C-1-A MSD**

**Matrix: Water**

**Analysis Batch: 109924**

**Client Sample ID: Matrix Spike Duplicate**

**Prep Type: Total/NA**

**Prep Batch: 109690**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier				Limits		
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		50.2	52.6		ng/L		105	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	13		50.2	60.5		ng/L		94	70 - 130	1	30
Perfluorododecanoic acid (PFDoA)	<2.0		50.2	47.4		ng/L		94	70 - 130	4	30
Perfluorooctanoic acid (PFOA)	4.9		50.2	55.9		ng/L		102	70 - 130	5	30
Perfluorodecanoic acid (PFDA)	<2.0		50.2	50.3		ng/L		100	70 - 130	5	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		50.2	55.4		ng/L		107	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	2.5		50.2	46.3		ng/L		92	70 - 130	3	30
Perfluoroheptanoic acid (PFHpA)	2.1		50.2	46.9		ng/L		89	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		50.2	50.9		ng/L		101	70 - 130	3	30
Perfluorotetradecanoic acid (PFTA)	<2.0		50.2	45.4		ng/L		90	70 - 130	4	30
Perfluorotridecanoic acid (PFTrDA)	<2.0		50.2	45.0		ng/L		90	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		50.2	56.4		ng/L		112	70 - 130	7	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		50.2	52.3		ng/L		104	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		50.2	45.4		ng/L		90	70 - 130	6	30
<b>Surrogate</b>	<b>MSD</b>	<b>MSD</b>	<b>Qualifier</b>	<b>Limits</b>							
d5-NEtFOSAA	111			70 - 130							
13C2 PFHxA	100			70 - 130							
13C2 PFDA	107			70 - 130							
13C3-GenX	105			70 - 130							

# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## LCMS

### Prep Batch: 109544

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
MBL 380-109544/22-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-109544/24-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-109544/23-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-113907-1 MS	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
380-113907-1 MSD	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	

### Analysis Batch: 109606

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	109544
MBL 380-109544/22-A	Method Blank	Total/NA	Water	537.1	109544
LCS 380-109544/24-A	Lab Control Sample	Total/NA	Water	537.1	109544
MRL 380-109544/23-A	Lab Control Sample	Total/NA	Water	537.1	109544
380-113907-1 MS	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	109544
380-113907-1 MSD	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	109544

### Prep Batch: 109690

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
MBL 380-109690/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-109690/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-109690/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-113655-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-113655-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

### Analysis Batch: 109924

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	109690
MBL 380-109690/20-A	Method Blank	Total/NA	Water	537.1	109690
LCS 380-109690/22-A	Lab Control Sample	Total/NA	Water	537.1	109690
MRL 380-109690/21-A	Lab Control Sample	Total/NA	Water	537.1	109690
380-113655-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	109690
380-113655-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	109690

### Prep Batch: 110608

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
MBL 380-110608/19-A	Method Blank	Total/NA	Water	533	
LCS 380-110608/21-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-110608/20-A	Lab Control Sample	Total/NA	Water	533	
380-113854-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-113854-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

### Analysis Batch: 110795

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113907-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	110608
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	110608
MBL 380-110608/19-A	Method Blank	Total/NA	Water	533	110608
LCS 380-110608/21-A	Lab Control Sample	Total/NA	Water	533	110608
MRL 380-110608/20-A	Lab Control Sample	Total/NA	Water	533	110608

Eurofins Eaton Analytical Pomona

# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## LCMS (Continued)

### Analysis Batch: 110795 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-113854-B-1-A MS	Matrix Spike	Total/NA	Water	533	110608
380-113854-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	110608

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

# Lab Chronicle

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-113907-1  
 SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-1**

**Date Collected: 09/17/24 10:00**

**Matrix: Water**

**Date Received: 09/19/24 09:59**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			110608	N8NE	EA POM	09/27/24 10:53
Total/NA	Analysis	533		1	110795	M7ML	EA POM	09/29/24 11:34
Total/NA	Prep	537.1 DW			109544	E9PK	EA POM	09/21/24 10:10
Total/NA	Analysis	537.1		1	109606	M7ML	EA POM	09/22/24 14:45

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-113907-2**

**Date Collected: 09/17/24 10:00**

**Matrix: Water**

**Date Received: 09/19/24 09:59**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			110608	N8NE	EA POM	09/27/24 10:53
Total/NA	Analysis	533		1	110795	M7ML	EA POM	09/29/24 11:45
Total/NA	Prep	537.1 DW			109690	U7RS	EA POM	09/23/24 09:19
Total/NA	Analysis	537.1		1	109924	M7ML	EA POM	09/24/24 17:34

**Laboratory References:**

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

## Laboratory: Eurofins Eaton Analytical Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-25

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17



# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-113907-1  
SDG: Weekly PFAS

---

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
380-113907-1	HALAWA SHAFT VIEWING POOL	Water	09/17/24 10:00	09/19/24 09:59
380-113907-2	FB: HALAWA SHAFT VIEWING POOL	Water	09/17/24 10:00	09/19/24 09:59

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17



ORIGIN ID:HIKA (808) 748-5040  
BWS CHEMLAB  
HONOLULU BOARD OF WATER SUPPLY  
630 S BERETANIA ST  
CHEMICAL LABORATORY  
HONOLULU, HI 96843  
UNITED STATES US

SHIP DATE: 18SEP24  
ACTWGHT 60.00 LB  
CAD: 258050552/INET4535  
DIMS 20x16x38 IN  
BILL RECIPIENT

TO **EUROFINS RECEIVING DEPARTMENT**  
**EUROFINS DRINKING WATER TESTING**  
**941 CORPORATE CENTER DR**

**POMONA CA 91768**

(626) 386-1100 REF: 583J24EF9/9AEC3



1 of 2  
TRK# 7786 7384 2213  
# MASTER # 91768  
**WM ONTA** CA-US ONT  
THU - 19 SEP 10:30A  
PRIORITY OVERNIGHT



After printing this label  
1 Fold the printed page along the horizontal line  
2 Place label in shipping pouch and affix it to your shipment

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-113907-1

SDG Number: Weekly PFAS

**Login Number: 113907**

**List Source: Eurofins Eaton Analytical Pomona**

**List Number: 1**

**Creator: Edrosa, Rey**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

