

750 Royal Oaks Drive, Suite 100  
Monrovia, California 91016-3629  
Tel: (626) 386-1100  
Fax: (866) 988-3757  
1 800 566 LABS (1 800 566 5227)

## Laboratory Report

for

Honolulu Board of Water Supply  
630 South Beretania Street  
Public Service Bldg.” Room 308  
Honolulu, HI 96843  
Attention: Erwin Kawata  
Fax: 808-550-5018



Utah ELCP CA00006

DEB: Debbie L Frank  
Project Manager

Report: 979938  
Project: RED-HILL  
Group: Red-Hill Expanded List (Albuquerque+)

\* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

\* Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.

\* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

\* Test results relate only to the sample(s) tested.

\* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

\* This report shall not be reproduced except in full, without the written approval of the laboratory.

\* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

## STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	NE-OS-21-13
Arkansas	CA00006	Nevada	CA00006
California	2813	New Hampshire *	2959
Colorado	CA00006	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	CA00006
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	21-008R	Ohio - 537.1	87786
Hawaii	CA00006	Oregon *	4034
Idaho	CA00006	Pennsylvania *	68-00565
Illinois	200033	Puerto Rico	CA00006
Indiana	C-CA-01	Rhode Island	LAO00326
Iowa – Asbestos	413	South Carolina	87016
Kansas *	E-10268	South Dakota	CA11320
Kentucky	90107	Tennessee	TN02839
Louisiana *	LA008	Texas *	T104704230-20-18
Maine	CA00006	Utah (Primary AB) *	CA00006
Maryland	224	Vermont	VT0114
Marianas Islands	MP0004	Virginia *	460260
Massachusetts	M-CA006	Washington	C838
Michigan	9906	EPA Region 5	CA00006
Mississippi	CA00006	Los Angeles County Sanitation Districts	10264

\* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2917 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

Test(s)	Method(s)	Potable Water *	Waste Water	Test(s)	Method(s)	Potable Water *	Waste Water
Enterococci	Enterolert	x	x	Gross Alpha coprecipitation	SM 7110 C	x	x
Escherichia coli (Enumeration)	SM 9221 B.1 SM 9221 F	x		Hardness	SM 2340 B	x	x
Fecal Coliform (P/A and Enumeration)	SM 9221 C (MTF/EC), SM 9221 E (MTF/EC)	x	x	Hexavalent Chromium	EPA 218.6,	x	x
Fecal Streptococci and Enterococci	SM 9230 B	x	x	Hexavalent Chromium	EPA 218.7,	x	
Heterotrophic Bacteria	SM 9215 B	x		Hexavalent Chromium	SM 3500-Cr B		x
Legionella	Legiolert®	x		Inorganic Anions and DBPs	EPA 300.0	x	x
Pseudomonas aeruginosa	Idexx Pseudalart	x		Norganic Anions and DBPs	EPA 300.1	x	
Total Coliform (P/A and Enumeration)	SM 9221A, SM 9221B, SM 9221 C	x	x	Kjeldahl Nitrogen	EPA 351.2		x
Total Coliform, Total Coliform with Chlorine Present	SM 9221 B	x	x	Metals	EPA 200.7, EPA200.8	x	x
Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure)	SM 9223	x		Nitrosamines	EEA-Agilent 521.1 (GCMS-24250)	x	
Total Microcystins and Nodularins	EPA 546	X		Nitrate/Nitrite Nitrogen	EPA 353.2	x	x
Yeast and Mold	SM 9610	x		Odor	SM2150B	x	
1,2,3-Trichloropropane (TCP) at 5 PPT	CA SRL 524M-TCP	x		Organohalide Pesticides and PCB	EPA 505	x	
1,4-Dioxane	EPA 522	x		Ortho Phosphate	SM 4500P E	x	
2,3,7,8-TCDD	Modified EPA 1613 B	x		Oxyhalides Disinfection Byproducts	EPA 317.0	x	
Acrylamide	+ LCMS 2440)	x		Perchlorate	EPA 331.0	x	
Algal Toxins/Microcystin	+ LCMS 3570	x		Perchlorate (Low and High Levels)	EPA 314.0	x	
Alkalinity	SM 2320B	x	x	Perfluorinated Alkyl Acids	EPA 533, EPA 537, EPA 537.1	x	
Ammonia	EPA 350.1, SM 4500-NH3 H		x	PPCP and EDC	+ LCMS-2443	x	
Asbestos	EPA 100.2	x	x	pH	EPA 150.1 SM 4500-H+ B	x	x
Bicarbonate Alkalinity as HCO3	SM 2330 B	x	x	Phenolics – Low Level	+WC 2493 (EPA 420.2 and EPA 420.4 MOD)	x	x
BOD/CBOD	SM 5210 B		x	Phenylurea Pesticides/Herbicides	+ LCMS-2448	x	
Bromate	+ LCMS- 2447	x		Radium-226, Radium-228	GA Tech (Rad-2374)	x	
Carbonate as CO3	SM 2330 B	x	x	Radon-222	SM 7500RN	x	
Carbonyls	EPA 556	x	x	Residue (Filterable)	SM 2540C	x	x
Chemical Oxygen Demand	EPA 410.4, SM 5220D		x	Residue (Non-Filterable)	SM 2540D		x
Chlorinated Acids	EPA 515.4	x		Residue (Total)	SM 2540B		x
Chlorine Dioxide	Palin Test Chlordio X Plus, SM 4500-CLO2 D	x		Residue (Volatile)	EPA 160.4		x
Chlorine, Free, Combined, Total Residual, Chloramines	SM 4500-Cl G	x		Semi-Volatile Compounds	EPA 525.2	x	
Color	SM2120B	x		Silica	SM 4500-SiO2 C	x	x
Conductivity	EPA 120.1, SM 2510B	x	x	Sulfide	SM 4500-S D		x
Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated	SM 2330 B	x		Sulfite	SM 4500-SO3 B	x	x
Cyanide (Amenable)	SM 4500-CN G	x	x	Surfactants	SM 5540C	x	x
Cyanide (Free)	SM 4500CN F	x	x	Taste and Odor	SM 6040 E	x	
Cyanide (Total)	EPA 335.4	x	x	Total Organic Carbon	SM 5310 C	x	x
Cyanogen Chloride (Screen)	+ 335 Mod (WC-24467)	x		Total Phenols	EPA 420.1		x
Diquat and Paraquat	EPA 549.2	x		Total Phenols	EPA 420.4	x	x
DBP and HAA	SM 6251 B	x		Triazine Pesticides and their Degradates	+ LCMS-3617	x	
Dissolved Organic Carbon	SM 5310 C	x		Turbidity	EPA 180.1	x	x
Dissolved Oxygen	SM 4500-O G		x	Uranium by ICP/MS	EPA 200.8	x	
EDB/DCBP/TCP	EPA 504.1	x		UV 254 Organic Constituents	SM 5910B	x	
EDB/DBCP and Disinfection Byproducts	EPA 551.1	x		VOCs	EPA 524.2	x	
EDTA and NTA	+ WC-2454	x		VOCs	+ (GCMS 2412) by EPA 524.2 modified	x	
Endothall	EPA 548.1, +(LCMS-2445)	x					
Fluoride	SM 4500F C	x	x				
Glyphosate	EPA 547	x					
Glyphosate and AMPA	+ LCMS-3618	x					
Gross Alpha and Gross Beta	EPA 900.0	x	x				

(\* ) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+ ) In-House Method

### Acknowledgement of Samples Received

Addr: **Honolulu Board of Water Supply**  
 630 South Beretania Street  
 Public Service Bldg." Room 308  
 Honolulu, HI 96843

Attn: Erwin Kawata  
 Phone: 808-748-5091

Client ID: HONOLULU

Folder #: 979938

Project: RED-HILL

Sample Group: Red-Hill Expanded List  
 (Albuquerque+)

Project Manager: Debbie L Frank

Phone: (626) 386-1149

PO #: C20525101 exp 05312023

The following samples were received from you on **January 12, 2022** at **1425**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
<u>202201120751</u>	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	01/11/2022 0818
	(SUB)Gas Fraction Hydrocarbons      TPH 8015 Diesel and Motor Oil      TPH 8015 Jet Fuel 5	
	TPH 8015 Jef Fuel 8	
<u>202201120752</u>	TRAVEL BLANK::AIEA GULCH WELLS PUMP 2 -331-202-TP072	01/11/2022 0818
	(SUB)Gas Fraction Hydrocarbons	

### Test Description

# CHAIN OF CUSTODY RECORD

979938

EUROFINS EATON ANALYTICAL USE ONLY:

750 Royal Oaks Drive, Suite 100  
 Monrovia, CA 91016-3629  
 Phone: 626 386 1100  
 Fax: 626 386 1101  
 800 566 LABS (800 566 5227)

LOGIN COMMENTS: \_\_\_\_\_ SAMPLES CHECKED AGAINST COC BY: R

SAMPLES LOGGED IN BY: R SAMPLES REC'D DAY OF COLLECTION?  (check for yes)

SAMPLE TEMP RECEIVED AT: \_\_\_\_\_ °C (Compliance: 4 ± 2 °C)  
 Colton / No. California / Arizona  
 Monrovia 3.0 °C (Compliance: 4 ± 2 °C)

CONDITION OF BLUE ICE: Frozen  Partially Frozen \_\_\_\_\_ Thawed \_\_\_\_\_ Wet Ice \_\_\_\_\_ No Ice \_\_\_\_\_


METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: \_\_\_\_\_

TO BE COMPLETED BY SAMPLER: \_\_\_\_\_ (check for yes)

COMPANY/AGENCY NAME:		PROJECT CODE:	COMPLIANCE SAMPLES	NON-COMPLIANCE SAMPLES	(check for yes)	
BWS HONOLULU		Red Hill	<input type="checkbox"/>	<input checked="" type="checkbox"/>		
EEA CLIENT CODE:		SAMPLE GROUP:	Type of samples (circle one): ROUTINE SPECIAL CONFIRMATION (eg. SDWA, Phase V, NPDES, FDA,...)			
TAT requested: rush by adv notice only		STD ___ 1 wk ___ X ___ 3 day ___ 2 day ___ 1 day ___	SEE ATTACHED BOTTLE ORDER FOR ANALYSES (check for yes), <u>OR</u>			
SAMPLE DATE	SAMPLE TIME	SAMPLE ID	8015 Diesel & Motor Oil	8015 Gac. C	8015 Gac. C TB	SAMPLER COMMENTS
01/11/22	0818	Aiea Gulch Wells Pump 2	X	X	X	
						Temp Blank: <u>6</u> °C

\* MATRIX TYPES: RSW = Raw Surface Water CFW = Chlor(am)inated Finished Water SEAW = Sea Water BW = Bottled Water SO = Soil  
 RGW = Raw Ground Water FW = Other Finished Water WW = Waste Water SW = Storm Water SL = Sludge

SIGNED BY:	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
RELINQUISHED BY:	<u>Olaf Happe</u>	Olaf Happe	Honolulu Board of Water Supply	1/11/2022	0818
RECEIVED BY:	<u>Olaf Happe</u>	Olaf Happe	Honolulu Board of Water Supply	1/11/2022	1100
RELINQUISHED BY:	<u>Selved...</u>		<u>ben</u>	1/12/22	1425
RECEIVED BY:					



Eaton Analytical

# INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number:

979935

### SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

### SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 4311 (Observation = 3.2 °C) (Corr.Factor = -0.2 °C) (Final = 3.0 °C)

TYPE OF ICE: Real  Synthetic  No Ice  CONDITION OF ICE: Frozen  Partially Frozen  Thawed  N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx UPS / DHL / Area Fast / Top Line / Other: \_\_\_\_\_

### Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr.Factor = _____ °C) (Final = _____ °C)

4 Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check. Manufacturer: \_\_\_\_\_ Lot Number: \_\_\_\_\_ pH strip type: 0 - 14 or \_\_\_\_\_ Expiration Date \_\_\_\_\_ Results: \_\_\_\_\_

6) Chlorine check. Manufacturer: Sansafe. Lot No.: \_\_\_\_\_ Expiration Date: \_\_\_\_\_ Results \_\_\_\_\_

7) VOA and Radon

No Samples with Headspace:

Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 515-4, HAA(6251,552), 505, SPME, @CH, 532LCMS, 556, 536, Anatoxin, LCMS methods using 40 ml vials, International clients:

Samp ID	Bottle #	None/<6 mm	>6mm	Test	Samp ID	Bottle #	None/<6 mm	>6mm	Test

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): \_\_\_\_\_

RECEIVED BY: [Signature] SIGNATURE [Signature] PRINT NAME Schadon COMPANY/TITLE Eurofins Eaton Analytical DATE 1/12/22 TIME 1425

SAMPLES CHECKED AGAINST COC BY: \_\_\_\_\_ SIGNATURE \_\_\_\_\_ PRINT NAME \_\_\_\_\_ COMPANY/TITLE Eurofins Eaton Analytical DATE \_\_\_\_\_ TIME \_\_\_\_\_

750 Royal Oaks Drive, Suite 100  
 Monrovia, California 91016-3629  
 (626) 386-1100 FAX (866) 988-3757

Created Date & Time: 12/10/2021 8:11:58PM

Kit #: 307677

Client ID: HONOLULU

Created By: - [AutoGenerated]  
 Deliver By: 12/29/2021  
 STG: Bottle Orders

Project Code: RED-HILL Bottle Orders  
 Group Name: Red-Hill Expanded List (Albuquerque+)  
 PO#/JOB#: C20525101 exp 05312023  
 Description: AIEA GULCH WELLS PUMP 2 - E

Ice Type: G  
 Pre Registered

**Note: Sampler Please return this paper with your samples**

**Ship Sample Kits to**  
 Honolulu Board of Water Supply  
 630 South Beretania Street  
 Chemistry Lab  
 Honolulu, HI 96843  
 Attn: Ron Fenstermacher  
 Phone: 808-748-5841  
 Fax: 808-550-5572

**Send Report to**  
 Honolulu Board of Water Supply  
 630 South Beretania Street  
 Public Service Bldg." Room 308  
 Honolulu, HI 96843  
 Attn: Erwin Kawata  
 Phone: 808-748-5091  
 Fax: 808-550-5018

**Billing Address**  
 Honolulu Board of Water Supply  
 630 South Beretania Street  
 Public Service Bldg." Room 308  
 Honolulu, HI 96843  
 Attn: Erwin Kawata  
 Phone: 808-748-5091  
 Fax: 808-550-5018

# of Sample Tests	Bottle Qty - Type [preservative information]	Total	UN DOT #
1	<del>@625A-Physis C, @625BN-Physis-G, @625PAH-Physis-HGS-C - 6 - 1L amber glass [1 ml Thio 8%]</del>	8	
1	TPH 8015 Diesel and Motor Oil_C, TPH 8015 Jet Fuel 5_C, TPH 8015 Jet Fuel 8_C	9	
1	8015 Gas_C	3	
1	8015 Gas_C TB	2	
4	<del>@VOASDWA-C plus IICs-TBC</del>	9	UN1789
4	<del>@VOASDWA-C plus plus TICs-C</del>	3	UN1789
4	<del>@8048-Ethanol-Subbed</del>	4	

**Sum Bottles: 32**

**Comments**

1st MS/MSD  
 SITE ID:  
 331-202 AIEA GULCH WELLS PUMP 2  
 SAMPLER:  
 EIGHT 1 LITER AMBER GLASS BOTTLES FOR 625 SERIES AND NINE 1 LITER AMBER GLASS BOTTLES FOR TPH 8015 SERIES. THIS IS A MSMSD SITE.  
 SHIPPING:  
 Travel Blanks - TBA/MTBE, VOASDWA - Prepare TBs in the VOA LAB.  
 Label Cooler on TOP and right below both Handles with Site description of contents ( use extra Container Labels)  
 ASM: Be sure to coordinate Follow-up as needed for any new detections in Field samples.  
 Acetone - follow-ups need to use EPA 624

ORIGIN ID:HIKA (808) 748-5840  
BWS CHEM LAB  
HONOLULU BOARD OF WATER SUPPLY  
630 S. BERETANIA ST.  
CHEMICAL LABORATORY  
HONOLULU, HI 96843  
UNITED STATES US

SHIP DATE: 11JAN22  
ACTWGT: 50.00 LB  
CAD: 100205419/NET74400  
BILL RECIPIENT


TO C CHUCK

EUROFINS EATON ANALYTICAL, INC  
750 ROYAL OAKS DR  
SUITE 100  
MONROVIA CA 91016

REF: (626) 386-1178  
INV:  
PO:

DEPT:

56DJ201EF/FE4A



4 of 8

MPS# 7757 2326 9028  
Mstr# 7757 2326 6474

0263  
0201

WED - 12 JAN 10:30A  
PRIORITY OVERNIGHT

WZ WHPA

91016  
CA-US BUR



**After printing this label:**

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

**Warning:** Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.  
Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our ServiceGuide. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Tel: (626) 386-1100  
Fax: (866) 988-3757  
1 800 566 LABS (1 800 566 5227)

**Laboratory Comments**

**Report:** 979938  
**Project:** RED-HILL  
**Group:** Red-Hill Expanded List  
(Albuquerque+)

Honolulu Board of Water Supply  
Erwin Kawata  
630 South Beretania Street  
Public Service Bldg." Room 308  
Honolulu, HI 96843

---

**Folder Comments**

Results for TPH Gas, Diesel, Motor Oil, Jet Fuels are submitted by EMAX Laboratories

Tel: (626) 386-1100  
Fax: (866) 988-3757  
1 800 566 LABS (1 800 566 5227)

**Report:** 979938  
**Project:** RED-HILL  
**Group:** Red-Hill Expanded List  
(Albuquerque+)

**Honolulu Board of Water Supply**  
Erwin Kawata  
630 South Beretania Street  
Public Service Bldg." Room 308  
Honolulu, HI 96843

Samples Received on:  
01/12/2022 1425

---

Analyzed	Analyte	Sample ID	Result	HI Limit	Units	MRL
----------	---------	-----------	--------	----------	-------	-----

---

Tel: (626) 386-1100  
 Fax: (626) 988-3757  
 1 800 566 LABS (1 800 566 5227)

Laboratory Data

**Report:** 979938  
**Project:** RED-HILL  
**Group:** Red-Hill Expanded List  
 (Albuquerque+)

**Honolulu Board of Water Supply**  
 Erwin Kawata  
 630 South Beretania Street  
 Public Service Bldg." Room 308  
 Honolulu, HI 96843

Samples Received on:  
 01/12/2022 1425

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
<b><u>AIEA GULCH WELLS PUMP 2 (331-202-TP072) (202201120751)</u></b>						<b>Sampled on 01/11/2022 0818</b>			
<b>SW 8015B - (SUB)Gas Fraction Hydrocarbons</b>									
01/13/22	01/13/22 15:30			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1
<b>SW 8015B - TPH 8015 Diesel and Motor Oil</b>									
01/13/22	01/14/22 22:08			(SW 8015B)	TPH Diesel	ND	mg/L	0.028	1
01/13/22	01/14/22 22:08			(SW 8015B)	TPH Motor Oil	ND	mg/L	0.055	1
<b>EPA 8015 - Jet Fuel 5 C8-C18</b>									
01/13/22	01/14/22 22:08			(EPA 8015)	Jet Fuel 5	ND	mg/L	0.055	1
<b>EPA 8015 - Jet Fuel 8 C8-C18</b>									
	01/14/22 22:08			(EPA 8015)	Jet Fuel 8	ND	mg/L	0.055	1
<b><u>TRAVEL BLANK::AIEA GULCH WELLS PUMP 2 -331-202-TP072 (202201120752)</u></b>						<b>Sampled on 01/11/2022 0818</b>			
<b>SW 8015B - (SUB)Gas Fraction Hydrocarbons</b>									
01/13/22	01/13/22 16:07			(SW 8015B)	(SUB)Gas Fraction Hydrocarbons	ND	mg/L	0.02	1

Rounding on totals after summation.  
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.



3051 Fujita Street  
Torrance, CA 90505  
Tel: (310)-618-8889

Date: 01-26-2022  
EMAX Batch No.: 22A117

Attn: Jackie Contreras

Eurofins Eaton Analytical  
750 Royal Oaks Dr., Suite 100  
Monrovia, CA 91016-3629

Subject: Laboratory Report  
Project: 979938


-----  
Enclosed is the Laboratory report for samples received on 01/13/22.  
The data reported relate only to samples listed below :

Sample ID	Control #	Col Date	Matrix	Analysis
202201120751	A117-01	01/11/22	WATER	TPH GASOLINE TPH
202201120752	A117-02	01/11/22	WATER	TPH GASOLINE

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.


Sincerely yours,

  
Caspar J. Pang  
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912021-19  
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing  
California ELAP Accredited Certificate Number 2672



Eaton Analytical

Ship To:  
EMAX Laboratories, Inc.  
3051 Fujita St.  
Torrance, CA 90505

Phone: 310-618-8889 Fax: 310-618-0818

Folder #: 979938 Report Due: 01/17/2022

Sample ID: 202201120751 Client Sample ID for reference on: AIEA GULCH WELLS PUMP 2 (331-202-TP072)

Sample type: SW 8015B EPA 5030C (SUB)Gas Fraction Hydrocarbons

SW 8015B EPA 3550B TPH 8015 Diesel and Motor Oil

EPA 8015 EPA 8015 Jet Fuel 5 C8-C18

EPA 8015 EPA 8015 Jet Fuel 8 C8-C18

Sample ID: 202201120752 Client Sample ID for reference on: TRAVEL BLANK: AIEA GULCH WELLS PUMP 2 -331-202-TP072

Sample type: SW 8015B EPA 5030C (SUB)Gas Fraction Hydrocarbons

Method: SW 8015B EPA 5030C (SUB)Gas Fraction Hydrocarbons

Date: 1/13/2022

Submittal Form

\*REPORTING REQUIREMENTS: Do Not Combine Reports with any other samples submitted under different Folder Numbers! Report & Invoice must have the Folder # 979938 Job # 1000014

Report all quality control data according to Method. Include dates analyzed. Date extracted (if extracted) and Method reference on the report. Results must have Complete data & QC with Approval Signature.

Reports: Jackie Contreras Sub-Contracting Administrator  
EMAIL TO: Eaton-MonroviaSubContract@eurofins.com  
Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016  
Phone (626) 386-1165 Fax (626) 386-1122  
Invoices to: Eurofins Eaton Analytical, LLC  
Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the Specified State Certification # and Exp Date for requested tests + matrix.  
Samples from: HAWAII

3 day rush

Sample Date & Time Matrix 01/11/22 0818 DW Clip Code PWSID JLS

Sample Point ID: Facility ID: Static ID:

Sample Date & Time Matrix 01/11/22 0818 DW Clip Code PWSID JLS

Sample Point ID: Facility ID: Static ID:

Relinquished by: Xjn Sample Control Date 1/13/22 Time 1140  
Received by: [Signature] Sample Control Date 01/13/22 Time 11:40  
Relinquished by: [Signature] Sample Control Date \_\_\_\_\_ Time \_\_\_\_\_  
Received by: \_\_\_\_\_ Date \_\_\_\_\_ Time \_\_\_\_\_

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS  
An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Temp: 3.1°  
3 2.1°

Type of Delivery <input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others	Airbill / Tracking Number	ECN <u>22A117</u>
<input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery		Recipient <u>Jocelyne Solis-Ramos</u>
		Date <u>01/13/22</u> Time <u>1:40</u>

**COC INSPECTION**

<input checked="" type="checkbox"/> Client Name	<input checked="" type="checkbox"/> Client PM/FC	<input type="checkbox"/> Sampler Name	<input checked="" type="checkbox"/> Sampling Date/Time	<input checked="" type="checkbox"/> Sample ID	<input checked="" type="checkbox"/> Matrix
<input checked="" type="checkbox"/> Address	<input checked="" type="checkbox"/> Tel # / Fax #	<input type="checkbox"/> Courier Signature	<input checked="" type="checkbox"/> Analysis Required	<input type="checkbox"/> Preservative (if any)	<input checked="" type="checkbox"/> TAT
Safety Issues (if any)	<input type="checkbox"/> High concentrations expected	<input type="checkbox"/> From Superfund Site	<input type="checkbox"/> Rad screening required		

Note: \_\_\_\_\_

**PACKAGING INSPECTION**

Container	<input checked="" type="checkbox"/> Cooler	<input type="checkbox"/> Box	<input type="checkbox"/> Other
Condition	<input type="checkbox"/> Custody Seal	<input type="checkbox"/> Intact	<input type="checkbox"/> Damaged
Packaging	<input checked="" type="checkbox"/> Bubble Pack	<input type="checkbox"/> Styrofoam	<input type="checkbox"/> Popcorn
Temperatures (Cool, ≤6 °C but not frozen)	<input type="checkbox"/> Cooler 1 _____ °C	<input checked="" type="checkbox"/> Cooler 2 <u>3.1</u> °C	<input checked="" type="checkbox"/> Cooler 3 <u>2.1</u> °C
	<input type="checkbox"/> Cooler 6 _____ °C	<input type="checkbox"/> Cooler 7 _____ °C	<input type="checkbox"/> Cooler 8 _____ °C
Thermometer: <u>A - S/N 210191066 a 12/14</u>	<u>B - S/N 210271396</u>	<u>C - S/N 210271399</u>	<u>D - S/N _____</u>

Comments:  Temperature is out of range. PM was informed IMMEDIATELY.

Note: \_\_\_\_\_

**DISCREPANCIES**

LabSampleID	LabSampleContainerID	Code	ClientSample Label ID / Information	Corrective Action
<u>1</u>	<u>4-7</u>	<u>D2</u>	<u>Jet fuel B is not indicated on label</u>	<u>R8</u>

4/13/22

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time.

MB 1/14/22

**NOTES/OBSERVATIONS:**

**LEGEND:**

- Code Description-Sample Management**
- D1 Analysis is not indicated in \_\_\_\_\_
  - D2 Analysis mismatch COC vs label
  - D3 Sample ID mismatch COC vs label
  - D4 Sample ID is not indicated in \_\_\_\_\_
  - D5 Container -[improper] [leaking] [broken]
  - D6 Date/Time is not indicated in \_\_\_\_\_
  - D7 Date/Time mismatch COC vs label
  - D8 Sample listed in COC is not received
  - D9 Sample received is not listed in COC
  - D10 No initial/date on corrections in COC/label
  - D11 Container count mismatch COC vs received
  - D12 Container size mismatch COC vs received

- Code Description-Sample Management**
- D13 Out of Holding Time
  - D14 Bubble is >6mm
  - D15 No trip blank in cooler
  - D16 Preservation not indicated in \_\_\_\_\_
  - D17 Preservation mismatch COC vs label
  - D18 Insufficient chemical preservative
  - D19 Insufficient Sample
  - D20 No filtration info for dissolved analysis
  - D21 No sample for moisture determination
  - D22 \_\_\_\_\_
  - D23 \_\_\_\_\_
  - D24 \_\_\_\_\_

Continue to next page.

- Code Description-Sample Management**
- R1 Proceed as indicated in  COC  Label
  - R2 Refer to attached instruction
  - R3 Cancel the analysis
  - R4 Use vial with smallest bubble first
  - R5 Log-in with latest sampling date and time+1 min
  - R6 Adjust pH as necessary
  - R7 Filter and preserved as necessary
  - R8 Informed client
  - R9 \_\_\_\_\_
  - R10 \_\_\_\_\_
  - R11 \_\_\_\_\_
  - R12 \_\_\_\_\_

**REVIEWS:**

Sample Labeling Jocelyne Solis-Ramos  
 Date 01/13/22

SRF Capitla  
 Date 1/13/22

PM MB  
 Date 1/14/22

## REPORTING CONVENTIONS

### DATA QUALIFIERS:

Lab Qualifier	AFCEE Qualifier	Description
J	F	Indicates that the analyte is positively identified and the result is less than LOQ/RL but greater than LOD/MDL/DL.
N		Indicates presumptive evidence of a compound.
B	B	Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level.
E	J	Indicates that the result is above the maximum calibration range or estimated value.
*	*	Out of QC limit.

**Note:** The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

### ACRONYMS AND ABBREVIATIONS:

CRDL	Contract Required Detection Limit
RL	Reporting Limit
MRL	Method Reporting Limit
MDL	Method Detection Limit
DL	Detection Limit
LOD	Limit of Detection
LOQ	Limit of Quantitation
DO	Diluted out

### DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

979938

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22A117



CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 979938

SDG : 22A117

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

A total of two(2) water samples were received on 01/13/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VG39A07B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VG39A07L/VG39A07C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

No matrix QC sample was provided on this SDG. Gasoline was within MS QC limits in A100-01M/A100-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL
Project     : 979938
SDG NO.    : 22A117
Instrument ID : GCT039
=====

```

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	WATER		Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
				Analysis DateTime	Analysis DateTime					
MBLK1W	VG39A07B	1	NA	01/13/2205:12	01/13/2205:12	EA12028A	EA12025A	EA12025A	22VG39A07	Method Blank
LCS1W	VG39A07L	1	NA	01/13/2205:49	01/13/2205:49	EA12029A	EA12025A	EA12025A	22VG39A07	Lab Control Sample (LCS)
LCD1W	VG39A07C	1	NA	01/13/2206:25	01/13/2206:25	EA12030A	EA12025A	EA12025A	22VG39A07	LCS Duplicate
202201120751	A117-01	1	NA	01/13/2215:30	01/13/2215:30	EA12045A	EA12036A	EA12036A	22VG39A07	Field Sample
202201120752	A117-02	1	NA	01/13/2216:07	01/13/2216:07	EA12046A	EA12036A	EA12036A	22VG39A07	Field Sample

```

FN      - Filename
% Moist - Percent Moisture

```

# SAMPLE RESULTS

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/11/22 08:18
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 15:30
Sample ID   : 202201120751              Date Analyzed: 01/13/22 15:30
Lab Samp ID: A117-01                     Dilution Factor: 1
Lab File ID: EA12045A                    Matrix: WATER
Ext Btch ID: 22VG39A07                   % Moisture: NA
Calib. Ref.: EA12036A                    Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
GASOLINE	ND	0.020	0.010	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0329	0.0400	82	60-140

Notes:

Parameter H-C Range  
Gasoline C6-C10  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 5ml Final Volume : 5ml  
Prepared by : SCerva Analyzed by : SCerva

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

=====  
Client : EUROFINS EATON ANALYTICAL Date Collected: 01/11/22 08:18  
Project : 979938 Date Received: 01/13/22  
Batch No. : 22A117 Date Extracted: 01/13/22 16:07  
Sample ID : 202201120752 Date Analyzed: 01/13/22 16:07  
Lab Samp ID: A117-02 Dilution Factor: 1  
Lab File ID: EA12046A Matrix: WATER  
Ext Btch ID: 22VG39A07 % Moisture: NA  
Calib. Ref.: EA12036A Instrument ID: 39  
=====

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
GASOLINE	ND	0.020	0.010	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0333	0.0400	83	60-140

Notes:  
Parameter H-C Range  
Gasoline C6-C10  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 5ml Final Volume : 5ml  
Prepared by : SCerva Analyzed by : SCerva

# QC SUMMARIES

METHOD 5030B/8015B  
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/13/22 05:12
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 05:12
Sample ID   : MBLK1W                     Date Analyzed: 01/13/22 05:12
Lab Samp ID: VG39A07B                   Dilution Factor: 1
Lab File ID: EA12028A                   Matrix: WATER
Ext Btch ID: 22VG39A07                 % Moisture: NA
Calib. Ref.: EA12025A                   Instrument ID: 39
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)		
GASOLINE	ND	0.020	0.010		
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT	
Bromofluorobenzene	0.0328	0.0400	82	60-140	

Notes:  
Parameter H-C Range  
Gasoline C6-C10  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 5ml                      Final Volume : 5ml  
Prepared by : SCerva                      Analyzed by : SCerva

EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979938  
BATCH NO. : 22A117  
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : MBLK1W                             LCS1W         LCD1W
LAB SAMPLE ID : VG39A07B                         VG39A07L     VG39A07C
LAB FILE ID  : EA12028A                         EA12029A     EA12030A
DATE PREPARED : 01/13/22 05:12                 01/13/22 05:49 01/13/22 06:25
DATE ANALYZED : 01/13/22 05:12                 01/13/22 05:49 01/13/22 06:25
PREP BATCH   : 22VG39A07                       22VG39A07   22VG39A07
CALIBRATION REF: EA12025A                       EA12025A    EA12025A
=====

```

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.498	100	0.500	0.482	96	3	60-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0427	107	0.0400	0.0427	107	70-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate



EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979513  
BATCH NO. : 22A100  
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202201110305                       202201110305MS
LAB SAMPLE ID : A100-01                          A100-01S
LAB FILE ID  : EA12039A                          EA12040A
DATE PREPARED : 01/13/22 11:52                   01/13/22 12:28
DATE ANALYZED : 01/13/22 11:52                   01/13/22 12:28
PREP BATCH   : 22VG39A07                         22VG39A07
CALIBRATION REF: EA12036A                        EA12036A
=====
  
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.456	91	0.500	0.482	96	6	50-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0415	104	0.0400	0.0413	103	60-140

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

979938

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22A117

## CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 979938

SDG : 22A117

### METHOD 3520C/8015B TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 01/13/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

#### Holding Time

The sample was analyzed within the prescribed holding time.

#### Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

#### Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSA010WB - result was compliant to project requirement. Refer to sample result summary form for details.

#### Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for Diesel was within LCS QC limits in DSA010WL. Refer to LCS summary form for details.

#### Matrix QC Sample

No matrix QC sample was provided on this SDG. One(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22A116-01M/22A116-01S. Refer to Matrix QC summary form for details.

#### Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

#### Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

## CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 979938

SDG : 22A117

### METHOD 3520C/8015B PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 01/13/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

#### Holding Time

The sample was analyzed within the prescribed holding time.

#### Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

#### Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSA010WB - result was compliant to project requirement. Refer to sample result summary form for details.

#### Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for JP5 was within LCS QC limits in J5A010WL. Refer to LCS summary form for details.

#### Matrix QC Sample

No matrix QC sample was provided on this SDG. One(1) set of MS/MSD was analyzed. JP5 was within MS QC limits in 22A116-01M/22A116-01S. Refer to Matrix QC summary form for details.

#### Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

#### Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 979938

SDG : 22A117

METHOD 3520C/8015B  
PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 01/13/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSA010WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for JP8 was within LCS QC limits in J8A010WL. Refer to LCS summary form for details.

Matrix QC Sample

No matrix QC sample was provided on this SDG. One(1) set of MS/MSD was analyzed. JP8 was within MS QC limits in 22A116-01M/22A116-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL
Project     : 979938
SDG NO.    : 22A117
Instrument ID : D5
=====

```

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	WATER			Extraction DateTime	Sample Data FN	Calibration Prep. Data FN	Batch	Notes
				Analysis DateTime	Extraction DateTime	Sample Data FN					
MBLK1W	DSA010MB	1	NA	01/14/2218:53	01/13/2216:00	LA14012A	LA14004A	22DSA010W	22DSA010W	Method Blank	
LCS1W	DSA010WL	1	NA	01/14/2219:11	01/13/2216:00	LA14013A	LA14004A	22DSA010W	22DSA010W	Lab Control Sample (LCS)	
202201120751	A117-01	1	NA	01/14/2222:08	01/13/2216:00	LA14023A	LA14004A	22DSA010W	22DSA010W	Field Sample	

```

FN      - Filename
% Moist - Percent Moisture

```

LAB CHRONICLE  
PETROLEUM HYDROCARBONS BY EXTRACTION

SDG NO. : 22A117  
Instrument ID : D5

Client : EUROFINS EATON ANALYTICAL  
Project : 979938

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	WATER			Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
				Analysis DateTime	Extraction DateTime	Sample Data FN					
MBLK1W	DSA010WB	1	NA	01/14/2218:53	01/13/2216:00	LA14012A	LA14005A	22DSA010W	Method Blank		
LCS1W	J5A010WL	1	NA	01/14/2219:28	01/13/2216:00	LA14014A	LA14005A	22DSA010W	Lab Control Sample (LCS)		
202201120751	A117-01	1	NA	01/14/2222:08	01/13/2216:00	LA14023A	LA14005A	22DSA010W	Field Sample		

FN - Filename  
% Moist - Percent Moisture

LAB CHRONICLE  
PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL
Project     : 979938
SDG NO.    : 22A117
Instrument ID : D5
=====

```

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	WATER		Extraction Date/Time	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
				Analysis Date/Time	% Moist					
MBLK1W	DSA010WB	1	NA	01/14/2218:53	01/13/2216:00	LA14012A	LA14006A	22DSA010W	Method Blank	
LCS1W	J8A010WL	1	NA	01/14/2219:46	01/13/2216:00	LA14015A	LA14006A	22DSA010W	Lab Control Sample (LCS)	
202201120751	A117-01	1	NA	01/14/2222:08	01/13/2216:00	LA14023A	LA14006A	22DSA010W	Field Sample	

```

FN      - Filename
% Moist - Percent Moisture

```



# SAMPLE RESULTS

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/11/22 08:18
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 16:00
Sample ID   : 202201120751              Date Analyzed: 01/14/22 22:08
Lab Samp ID : 22A117-01                  Dilution Factor: 1
Lab File ID : LA14023A                   Matrix: WATER
Ext Btch ID : 22DSA010W                   % Moisture: NA
Calib. Ref.: LA14004A                    Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.028	0.014	
Motor Oil	ND	0.055	0.028	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.449	0.550	82	60-130
Hexacosane	0.111	0.138	81	60-130

Notes:  
Parameter H-C Range  
Diesel C10-C24  
Motor Oil C24-C36  
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
Sample Amount : 910ml Final Volume : 5ml  
Prepared by : JMuert Analyzed by : SDeeso

METHOD 3520C/8015B  
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/11/22 08:18
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 16:00
Sample ID   : 202201120751              Date Analyzed: 01/14/22 22:08
Lab Samp ID: 22A117-01                   Dilution Factor: 1
Lab File ID: LA14023A                    Matrix: WATER
Ext Btch ID: 22DSA010W                   % Moisture: NA
Calib. Ref.: LA14005A                    Instrument ID: D5
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
JP5	ND	0.055	0.028	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.449	0.550	82	60-130
Hexacosane	0.111	0.138	81	60-130

Notes:

RL : Reporting Limit  
 Parameter H-C Range  
 JP5 C8-C18  
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
 Sample Amount : 910ml Final Volume : 5ml  
 Prepared by : JMuert Analyzed by : SDeeso

METHOD 3520C/8015B  
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/11/22 08:18
Project    : 979938                      Date Received: 01/13/22
Batch No.  : 22A117                      Date Extracted: 01/13/22 16:00
Sample ID  : 202201120751               Date Analyzed: 01/14/22 22:08
Lab Samp ID: 22A117-01                  Dilution Factor: 1
Lab File ID: LA14023A                   Matrix: WATER
Ext Btch ID: 22DSA010W                  % Moisture: NA
Calib. Ref.: LA14006A                   Instrument ID: D5
=====
    
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP8	ND	0.055	0.028

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.449	0.550	82	60-130
Hexacosane	0.111	0.138	81	60-130

Notes:

RL : Reporting Limit  
 Parameter H-C Range  
 JP8 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 910ml                      Final Volume : 5ml  
 Prepared by : JMuert                        Analyzed by : SDeeso

# QC SUMMARIES

METHOD 3520C/8015B  
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/13/22 16:00
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 16:00
Sample ID   : MBLK1W                     Date Analyzed: 01/14/22 18:53
Lab Samp ID : DSA010WB                   Dilution Factor: 1
Lab File ID : LA14012A                   Matrix: WATER
Ext Btch ID : 22DSA010W                  % Moisture: NA
Calib. Ref. : LA14004A                   Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.025	0.012	
Motor Oil	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.386	0.500	77	60-130
Hexacosane	0.109	0.125	87	60-130

Notes:

Parameter H-C Range  
Diesel C10-C24  
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml                      Final Volume : 5ml  
Prepared by : JMuert                              Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979938  
BATCH NO. : 22A117  
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA  
DILUTION FACTOR: 1 1  
SAMPLE ID : MBLK1W LCS1W  
LAB SAMPLE ID : DSA010WB DSA010WL  
LAB FILE ID : LA14012A LA14013A  
DATE PREPARED : 01/13/22 16:00 01/13/22 16:00  
DATE ANALYZED : 01/14/22 18:53 01/14/22 19:11  
PREP BATCH : 22DSA010W 22DSA010W  
CALIBRATION REF: LA14004A LA14004A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
Diesel	ND	2.50	2.39	96	50-130

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
Bromobenzene	0.500	0.465	93	60-130
Hexacosane	0.125	0.118	94	60-130

MB: Method Blank sample LCS: Lab Control Sample

METHOD 3520C/8015B  
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/13/22 16:00
Project     : 979938                      Date Received: 01/13/22
Batch No.   : 22A117                      Date Extracted: 01/13/22 16:00
Sample ID   : MBLK1W                      Date Analyzed: 01/14/22 18:53
Lab Samp ID: DSA010WB                     Dilution Factor: 1
Lab File ID: LA14012A                     Matrix: WATER
Ext Btch ID: 22DSA010W                    % Moisture: NA
Calib. Ref.: LA14005A                     Instrument ID: D5
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
JP5	ND	0.050	0.025

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.386	0.500	77	60-130
Hexacosane	0.109	0.125	87	60-130

Notes:

RL : Reporting Limit  
 Parameter H-C Range  
 JP5 C8-C18  
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.  
 Sample Amount : 1000ml Final Volume : 5ml  
 Prepared by : JMuert Analyzed by : SDeeso



EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979938  
BATCH NO. : 22A117  
METHOD : 3520C/8015B

=====

MATRIX	: WATER	% MOISTURE:NA
DILUTION FACTOR:	1	1
SAMPLE ID	: MBLK1W	LCS1W
LAB SAMPLE ID	: DSA010WB	J5A010WL
LAB FILE ID	: LA14012A	LA14014A
DATE PREPARED	: 01/13/22 16:00	01/13/22 16:00
DATE ANALYZED	: 01/14/22 18:53	01/14/22 19:28
PREP BATCH	: 22DSA010W	22DSA010W
CALIBRATION REF:	LA14005A	LA14005A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----	-----
JP5	ND	2.50	2.12	85	30-160

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----
Bromobenzene	0.500	0.477	95	60-130
Hexacosane	0.125	0.115	92	60-130

=====

MB: Method Blank sample    LCS: Lab Control Sample

METHOD 3520C/8015B  
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 01/13/22 16:00
Project     : 979938                     Date Received: 01/13/22
Batch No.   : 22A117                     Date Extracted: 01/13/22 16:00
Sample ID   : MBLK1W                     Date Analyzed: 01/14/22 18:53
Lab Samp ID: DSA010WB                   Dilution Factor: 1
Lab File ID: LA14012A                   Matrix: WATER
Ext Btch ID: 22DSA010W                   % Moisture: NA
Calib. Ref.: LA14006A                   Instrument ID: D5
=====
    
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
JP8	ND	0.050	0.025	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.386	0.500	77	60-130
Hexacosane	0.109	0.125	87	60-130

Notes:

RL : Reporting Limit  
 Parameter H-C Range  
 JP8 C8-C18  
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml                      Final Volume : 5ml  
 Prepared by : JMuert                        Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA  
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979938  
BATCH NO. : 22A117  
METHOD : 3520C/8015B

=====

MATRIX	: WATER	% MOISTURE:NA
DILUTION FACTOR:	1	1
SAMPLE ID	: MBLK1W	LCS1W
LAB SAMPLE ID	: DSA010WB	J8A010WL
LAB FILE ID	: LA14012A	LA14015A
DATE PREPARED	: 01/13/22 16:00	01/13/22 16:00
DATE ANALYZED	: 01/14/22 18:53	01/14/22 19:46
PREP BATCH	: 22DSA010W	22DSA010W
CALIBRATION REF:	LA14006A	LA14006A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----	-----
JP8	ND	2.50	1.99	80	30-160

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----
Bromobenzene	0.500	0.519	104	60-130
Hexacosane	0.125	0.107	86	60-130

=====

MB: Method Blank sample    LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979946  
BATCH NO. : 22A116  
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202201120772                       202201120772MSD
LAB SAMPLE ID : 22A116-01                         22A116-01S
LAB FILE ID  : LA14016A                          LA14018A
DATE PREPARED : 01/13/22 16:00                   01/13/22 16:00
DATE ANALYZED : 01/14/22 20:03                   01/14/22 20:39
PREP BATCH   : 22DSA010W                         22DSA010W
CALIBRATION REF: LA14004A                       LA14004A
=====

```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Diesel	ND	2.95	2.66	90	2.95	2.70	92	1	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.590	0.441	75	0.590	0.482	82	60-130
Hexacosane	0.148	0.124	84	0.148	0.128	87	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979946  
BATCH NO. : 22A116  
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202201120772                       202201120772MSD
LAB SAMPLE ID : 22A116-01                         22A116-01S
LAB FILE ID  : LA14016A                           LA14020A
DATE PREPARED : 01/13/22 16:00                   01/13/22 16:00
DATE ANALYZED : 01/14/22 20:03                   01/14/22 21:15
PREP BATCH   : 22DSA010W                           22DSA010W
CALIBRATION REF: LA14005A                           LA14005A
  
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
JP5	ND	2.78	2.37	85	2.80	2.86	102	19	30-160	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.555	0.496	89	0.560	0.540	96	60-130
Hexacosane	0.139	0.120	86	0.140	0.126	90	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

EMAX QUALITY CONTROL DATA  
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL  
PROJECT : 979946  
BATCH NO. : 22A116  
METHOD : 3520C/8015B

MATRIX	: WATER		% MOISTURE:NA
DILUTION FACTOR:	1	1	1
SAMPLE ID	: 202201120772	202201120772MS	202201120772MSD
LAB SAMPLE ID	: 22A116-01	22A116-01M	22A116-01S
LAB FILE ID	: LA14016A	LA14021A	LA14022A
DATE PREPARED	: 01/13/22 16:00	01/13/22 16:00	01/13/22 16:00
DATE ANALYZED	: 01/14/22 20:03	01/14/22 21:33	01/14/22 21:50
PREP BATCH	: 22DSA010W	22DSA010W	22DSA010W
CALIBRATION REF:	LA14006A	LA14006A	LA14006A

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
JP8	ND	2.88	2.41	84	2.88	2.85	99	17	30-160	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromobenzene	0.575	0.556	97	0.575	0.634	110	60-130
Hexacosane	0.144	0.118	82	0.144	0.125	87	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate