

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

Honolulu Board of Water Supply
630 South Beretania Street
Public Service Bldg.” Room 308
Honolulu, HI 96843
Attention: Erwin Kawata
Fax: 808-550-5018



Utah ELCP CA00006

DEB: Debbie L Frank
Project Manager

Report: 987974
Project: RED-HILL
Group: Red-Hill Expanded List (Albuquerque+)

* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

* Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.

* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

* Test results relate only to the sample(s) tested.

* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

* This report shall not be reproduced except in full, without the written approval of the laboratory.

* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

STATE CERTIFICATION LIST

| State | Certification Number | State | Certification Number |
|------------------|----------------------|---|----------------------|
| Alabama | 41060 | Montana | Cert 0035 |
| Arizona | AZ0778 | Nebraska | NE-OS-21-13 |
| Arkansas | CA00006 | Nevada | CA00006 |
| California | 2813 | New Hampshire * | 2959 |
| Colorado | CA00006 | New Jersey * | CA 008 |
| Connecticut | PH-0107 | New Mexico | CA00006 |
| Delaware | CA 006 | New York * | 11320 |
| Florida * | E871024 | North Carolina | 06701 |
| Georgia | 947 | North Dakota | R-009 |
| Guam | 21-008R | Ohio - 537.1 | 87786 |
| Hawaii | CA00006 | Oregon * | 4034 |
| Idaho | CA00006 | Pennsylvania * | 68-00565 |
| Illinois | 200033 | Puerto Rico | CA00006 |
| Indiana | C-CA-01 | Rhode Island | LAO00326 |
| Iowa – Asbestos | 413 | South Carolina | 87016 |
| Kansas * | E-10268 | South Dakota | CA11320 |
| Kentucky | 90107 | Tennessee | TN02839 |
| Louisiana * | LA008 | Texas * | T104704230-20-18 |
| Maine | CA00006 | Utah (Primary AB) * | CA00006 |
| Maryland | 224 | Vermont | VT0114 |
| Marianas Islands | MP0004 | Virginia * | 460260 |
| Massachusetts | M-CA006 | Washington | C838 |
| Michigan | 9906 | EPA Region 5 | CA00006 |
| Mississippi | CA00006 | Los Angeles County Sanitation Districts | 10264 |

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2917 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

| Test(s) | Method(s) | Potable Water * | Waste Water | Test(s) | Method(s) | Potable Water * | Waste Water |
|---|--|-----------------|-------------|--|--|-----------------|-------------|
| Enterococci | Enterolert | x | x | Gross Alpha coprecipitation | SM 7110 C | x | x |
| <i>Escherichia coli</i> (Enumeration) | SM 9221 B.1 SM 9221 F | x | | Hardness | SM 2340 B | x | x |
| Fecal Coliform (P/A and Enumeration) | SM 9221 C (MTF/EC), SM 9221 E (MTF/EC) | x | x | Hexavalent Chromium | EPA 218.6, | x | x |
| Fecal Streptococci and Enterococci | SM 9230 B | x | x | Hexavalent Chromium | EPA 218.7, | x | |
| Heterotrophic Bacteria | SM 9215 B | x | | Hexavalent Chromium | SM 3500-Cr B | | x |
| Legionella | Legiolert® | x | | Inorganic Anions and DBPs | EPA 300.0 | x | x |
| <i>Pseudomonas aeruginosa</i> | Idexx Pseudalart | x | | Norganic Anions and DBPs | EPA 300.1 | x | |
| Total Coliform (P/A and Enumeration) | SM 9221A, SM 9221B, SM 9221 C | x | x | Kjeldahl Nitrogen | EPA 351.2 | | x |
| Total Coliform, Total Coliform with Chlorine Present | SM 9221 B | x | x | Metals | EPA 200.7, EPA200.8 | x | x |
| Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure) | SM 9223 | x | | Nitrosamines | EEA-Agilent 521.1 (GCMS-24250) | x | |
| Total Microcystins and Nodularins | EPA 546 | X | | Nitrate/Nitrite Nitrogen | EPA 353.2 | x | x |
| Yeast and Mold | SM 9610 | x | | Odor | SM2150B | x | |
| 1,2,3-Trichloropropane (TCP) at 5 PPT | CA SRL 524M-TCP | x | | Organohalide Pesticides and PCB | EPA 505 | x | |
| 1,4-Dioxane | EPA 522 | x | | Ortho Phosphate | SM 4500P E | x | |
| 2,3,7,8-TCDD | Modified EPA 1613 B | x | | Oxyhalides Disinfection Byproducts | EPA 317.0 | x | |
| Acrylamide | + LCMS 2440) | x | | Perchlorate | EPA 331.0 | x | |
| Algal Toxins/Microcystin | + LCMS 3570 | x | | Perchlorate (Low and High Levels) | EPA 314.0 | x | |
| Alkalinity | SM 2320B | x | x | Perfluorinated Alkyl Acids | EPA 533, EPA 537, EPA 537.1 | x | |
| Ammonia | EPA 350.1, SM 4500-NH3 H | | x | PPCP and EDC | + LCMS-2443 | x | |
| Asbestos | EPA 100.2 | x | x | pH | EPA 150.1 SM 4500-H+ B | x | x |
| Bicarbonate Alkalinity as HCO3 | SM 2330 B | x | x | Phenolics – Low Level | +WC 2493 (EPA 420.2 and EPA 420.4 MOD) | x | x |
| BOD/CBOD | SM 5210 B | | x | Phenylurea Pesticides/Herbicides | + LCMS-2448 | x | |
| Bromate | + LCMS- 2447 | x | | Radium-226, Radium-228 | GA Tech (Rad-2374) | x | |
| Carbonate as CO3 | SM 2330 B | x | x | Radon-222 | SM 7500RN | x | |
| Carbonyls | EPA 556 | x | x | Residue (Filterable) | SM 2540C | x | x |
| Chemical Oxygen Demand | EPA 410.4, SM 5220D | | x | Residue (Non-Filterable) | SM 2540D | | x |
| Chlorinated Acids | EPA 515.4 | x | | Residue (Total) | SM 2540B | | x |
| Chlorine Dioxide | Palin Test Chlordio X Plus, SM 4500-CLO2 D | x | | Residue (Volatile) | EPA 160.4 | | x |
| Chlorine, Free, Combined, Total Residual, Chloramines | SM 4500-Cl G | x | | Semi-Volatile Compounds | EPA 525.2 | x | |
| Color | SM2120B | x | | Silica | SM 4500-SiO2 C | x | x |
| Conductivity | EPA 120.1, SM 2510B | x | x | Sulfide | SM 4500-S D | | x |
| Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated | SM 2330 B | x | | Sulfite | SM 4500-SO3 B | x | x |
| Cyanide (Amenable) | SM 4500-CN G | x | x | Surfactants | SM 5540C | x | x |
| Cyanide (Free) | SM 4500CN F | x | x | Taste and Odor | SM 6040 E | x | |
| Cyanide (Total) | EPA 335.4 | x | x | Total Organic Carbon | SM 5310 C | x | x |
| Cyanogen Chloride (Screen) | + 335 Mod (WC-24467) | x | | Total Phenols | EPA 420.1 | | x |
| Diquat and Paraquat | EPA 549.2 | x | | Total Phenols | EPA 420.4 | x | x |
| DBP and HAA | SM 6251 B | x | | Triazine Pesticides and their Degradates | + LCMS-3617 | x | |
| Dissolved Organic Carbon | SM 5310 C | x | | Turbidity | EPA 180.1 | x | x |
| Dissolved Oxygen | SM 4500-O G | | x | Uranium by ICP/MS | EPA 200.8 | x | |
| EDB/DCBP/TCP | EPA 504.1 | x | | UV 254 Organic Constituents | SM 5910B | x | |
| EDB/DBCP and Disinfection Byproducts | EPA 551.1 | x | | VOCs | EPA 524.2 | x | |
| EDTA and NTA | + WC-2454 | x | | VOCs | + (GCMS 2412) by EPA 524.2 modified | x | |
| Endothall | EPA 548.1, +(LCMS-2445) | x | | | | | |
| Fluoride | SM 4500F C | x | x | | | | |
| Glyphosate | EPA 547 | x | | | | | |
| Glyphosate and AMPA | + LCMS-3618 | x | | | | | |
| Gross Alpha and Gross Beta | EPA 900.0 | x | x | | | | |

(*) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+) In-House Method

Acknowledgement of Samples Received

Addr: **Honolulu Board of Water Supply**
 630 South Beretania Street
 Public Service Bldg." Room 308
 Honolulu, HI 96843

Attn: Erwin Kawata
 Phone: 808-748-5091

Client ID: HONOLULU
 Folder #: 987974
 Project: RED-HILL
 Sample Group: Red-Hill Expanded List
 (Albuquerque+)
 Project Manager: Debbie L Frank
 Phone: (626) 386-1149
 PO #: C20525101 exp 05312023

The following samples were received from you on **February 16, 2022** at **1810**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

| Sample # | Sample ID | Sample Date |
|---------------------|---|-----------------|
| <u>202202161178</u> | HALAWA WELLS 2 (331-024-WL064) | 02/14/2022 1026 |
| | (SUB)Gas Fraction Hydrocarbons TPH 8015 Diesel and Motor Oil TPH 8015 Jet Fuel 5 TPH 8015 Jef Fuel 8 | |
| <u>202202161179</u> | TRAVEL BLANK::HALAWA WELLS 2 (331-024-WL064) | 02/14/2022 1026 |
| | (SUB)Gas Fraction Hydrocarbons | |

Test Description



Eaton Analytical

Kit Order for BOARD OF WATER SUPPLY, CITY AND COUNTY OF

Debbie L Frank is your Eurofins Eaton Analytical, LLC Service Manager

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
(626) 386-1100 FAX (866) 988-3757

Created Date & Time: 1/10/2022 12:06:38AM

Note: Sampler Please return this paper with your samples

Kit #: 310071

Client ID: HONOLULU

Created By: - [AutoGenerated]
Deliver By: 02/09/2022
STG: Bottle Orders

Project Code: RED-HILL Bottle Orders
Group Name: Red-Hill Expanded List (Albuquerque+)
PO#/JOB#: C20525101 exp 05312023
Description: HALAWA WELLS UNITS 1 & 2 - er

Ice Type: G
Pre Registered

Ship Sample Kits to
Honolulu Board of Water Supply
630 South Beretania Street
Chemistry Lab
Honolulu, HI 96843
Attn: Ron Fenstermacher
Phone: 808-748-5841
Fax: 808-550-5572

Send Report to
Honolulu Board of Water Supply
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843
Attn: Erwin Kawata
Phone: 808-748-5091
Fax: 808-550-5018

Billing Address
Honolulu Board of Water Supply
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843
Attn: Erwin Kawata
Phone: 808-748-5091
Fax: 808-550-5018

| # of | Sample Tests | Bottle Qty - Type [preservative information] | Total | UN DOT # |
|------|---|--|-------|----------|
| 1 | TPH 8015 Diesel and Motor Oil_C, TPH 8015 Jet Fuel 5_C, TPH 8015 Jet Fuel 8_C | 6 - 1L amber glass [1 ml Thio 8%] | 6 | |
| 1 | 8015 Gas_C | 3 - 40ml amber glass vial [1 drop Thio (8%)] | 3 | |
| 1 | @504MOD TB C, 8015 Gas_C TB | 2 - 40ml amber glass vial [1 drop Thio (8%) + H2O] | 2 | |
| 4 | @VOASDWA-G-plus-plus-HCl-TBC | 3 - 40ml amber glass vial [25mg-AA+ H2O+10-drop-1:1-HCl] | 3 | UN1789 |

Sum Bottles: 14

Sum Tests: 4


Comments

SITE ID:
HALAWA WELLS (331-206-TP065)

SAMPLER:
FOUR 1 LITER AMBER GLASS BOTTLES FOR 625 SERIES AND SIX 1 LITER AMBER GLASS BOTTLES FOR TPH 8015 SERIES.

SHIPPING:
Travel Blanks - TBA/MTBE, VOASDWA - Prepare TBs in the VOA LAB.
Label Cooler on TOP and right below both Handles with Site description of contents (use extra Container Labels)

ASM: Be sure to coordinate Follow-up as needed for any new detections in Field samples.
Acetone - follow-ups need to use EPA 624



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: 48774

SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / (N)

IR Gun ID = 630 (Observation = 4.1 °C) (Corr. Factor = -0.2 °C) (Final = 3.9 °C)

TYPE OF ICE: Real Synthetic No Ice CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / (FedEx) / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

| | |
|---|---|
| 1 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C) | 2 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C) |
| 3 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C) | 4 = (Observation = _____ °C) (Corr. Factor = _____ °C) (Final = _____ °C) |

- 4 Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)
- 5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____
- 6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

VOA and Radon Headspace: No Samples with Headspace: Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

| Samp ID | Bottle # | None/<6 mm | >6mm | Test | Samp ID | Bottle # | None/<6 mm | >6mm | Test |
|---------|----------|------------|------|------|---------|----------|------------|------|------|
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |
| | | | | | | | | | |

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): _____

| | | | | |
|-----------|-----------------|---------------------------|---------|------|
| SIGNATURE | PRINT NAME | COMPANY/TITLE | DATE | TIME |
| | Victor Pascasio | Eurofins Eaton Analytical | 2.16.22 | 1550 |
| SIGNATURE | PRINT NAME | COMPANY/TITLE | DATE | TIME |
| | | Eurofins Eaton Analytical | | |

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 987974
Project: RED-HILL
Group: Red-Hill Expanded List
(Albuquerque+)

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Folder Comments

Results for Ethanol, TPH Gas, Diesel, Motor Oil and Jet Fuels are submitted by Emax Laboratories



Eaton Analytical

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Hits

Report: 987974
Project: RED-HILL
Group: Red-Hill Expanded List
(Albuquerque+)

Honolulu Board of Water Supply
Erwin Kawata
630 South Beretania Street
Public Service Bldg." Room 308
Honolulu, HI 96843

Samples Received on:
02/16/2022 1810

| Analyzed | Analyte | Sample ID | Result | HI Limit | Units | MRL |
|----------|---------|-----------|--------|----------|-------|-----|
|----------|---------|-----------|--------|----------|-------|-----|

SUMMARY OF POSITIVE DATA ONLY

Tel: (626) 386-1100
 Fax: (866) 988-3757
 1 800 566 LABS (1 800 566 5227)

Report: 987974
Project: RED-HILL
Group: Red-Hill Expanded List
 (Albuquerque+)

Honolulu Board of Water Supply
 Erwin Kawata
 630 South Beretania Street
 Public Service Bldg." Room 308
 Honolulu, HI 96843

Samples Received on:
 02/16/2022 1810

| Prepped | Analyzed | Prep Batch | Analytical Batch | Method | Analyte | Result | Units | MRL | Dilution |
|---|----------------|------------|------------------|------------|--------------------------------|-----------------------------------|-------|-------|----------|
| <u>HALAWA WELLS 2 (331-024-WL064) (202202161178)</u> | | | | | | Sampled on 02/14/2022 1026 | | | |
| SW 8015B - (SUB)Gas Fraction Hydrocarbons | | | | | | | | | |
| 02/18/22 | 02/18/22 01:22 | | | (SW 8015B) | (SUB)Gas Fraction Hydrocarbons | ND | mg/L | 0.02 | 1 |
| SW 8015B - TPH 8015 Diesel and Motor Oil | | | | | | | | | |
| 02/21/22 | 02/23/22 01:33 | | | (SW 8015B) | TPH Diesel | ND | mg/L | 0.026 | 1 |
| 02/21/22 | 02/23/22 01:33 | | | (SW 8015B) | TPH Motor Oil | ND | mg/L | 0.052 | 1 |
| EPA 8015 - Jet Fuel 5 C8-C18 | | | | | | | | | |
| 02/21/22 | 02/23/22 01:33 | | | (EPA 8015) | Jet Fuel 5 | ND | mg/L | 0.052 | 1 |
| EPA 8015 - Jet Fuel 8 C8-C18 | | | | | | | | | |
| | 02/23/22 01:33 | | | (EPA 8015) | Jet Fuel 8 | ND | mg/L | 0.052 | 1 |
| <u>TRAVEL BLANK::HALAWA WELLS 2 (331-024-WL064) (202202161179)</u> | | | | | | Sampled on 02/14/2022 1026 | | | |
| SW 8015B - (SUB)Gas Fraction Hydrocarbons | | | | | | | | | |
| 02/18/22 | 02/18/22 01:58 | | | (SW 8015B) | (SUB)Gas Fraction Hydrocarbons | ND | mg/L | 0.02 | 1 |

Rounding on totals after summation.
 (c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.



3051 Fujita Street
Torrance, CA 90505
Tel: (310)-618-8889

Date: 03-08-2022
EMAX Batch No.: 22B180

Attn: Jackie Contreras

Eurofins Eaton Analytical
750 Royal Oaks Dr., Suite 100
Monrovia, CA 91016-3629

Subject: Laboratory Report
Project: 987974

Enclosed is the Laboratory report for samples received on 02/17/22.
The data reported relate only to samples listed below :

| Sample ID | Control # | Col Date | Matrix | Analysis |
|--------------|-----------|----------|--------|---------------------|
| 202202161178 | B180-01 | 02/14/22 | WATER | TPH GASOLINE TPH |
| 202202161179 | B180-02 | 02/14/22 | WATER | TPH GASOLINE |

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.

Sincerely yours,


Caspar J. Pang
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912021-19
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing
California ELAP Accredited Certificate Number 2672



Eaton Analytical

Ship To:
EMAX Laboratories, Inc.
3051 Fujita St.
Torrance, CA 90505

Phone: 310-618-8889 Fax: 310-618-0818

Folder #: 987974 Report Due: 02/23/2022

Sample ID: 202202161178 Client Sample ID for reference onl
HALAWA WELLS 2 (331-024-WL064)

Sample type: Sample Event: Analysis Requested

| Method | Prep Method | Analysis Requested |
|----------|-------------|--------------------------------|
| SW 8015B | EPA 5030C | (SUB)Gas Fraction Hydrocarbons |
| SW 8015B | EPA 3550B | TPH 8015 Diesel and Motor Oil |
| EPA 8015 | EPA 8015 | Jet Fuel 5 C8-C18 |
| EPA 8015 | | Jet Fuel 8 C8-C18 |

Sample ID: 202202161179 Client Sample ID for reference onl
TRAVEL BLANK: HALAWA WELLS 2 (331-024-WL064)

| Method | Prep Method | Analysis Requested |
|----------|-------------|--------------------------------|
| SW 8015B | EPA 5030C | (SUB)Gas Fraction Hydrocarbons |

Relinquished by: *[Signature]* Sample Control
 Received by: *[Signature]*
 Relinquished by: Sample Control
 Received by:

Date: 02/17/22 Time: 12:14
 Date: 02/17/22 Time: 12:14
 Date: _____ Time: _____
 Date: _____ Time: _____

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS
 An Acknowledgement of Receipt is requested to attain Jackie Contreras
 Temp. 2.4/1.9, 1.9/1.4, 2.6/2.1

Submittal Form 22B180 Date: 2/17/2022
 *REPORTING REQUIREMENTS: Do Not Combine Reports with any other samples submitted under different Folder Numbers!
 Report & Invoice must have the Folder# 987974 Job # 1000014
 Report all quality control data according to Method. Include dates analyzed. Date extracted (if extracted) and Method reference on the report.
 Results must have Complete data & QC with Approval Signature.

Reports: Jackie Contreras Sub-Contracting Administrator
 EMAIL TO: Eaton-MonroviaSubContract@eurofinset.com
 Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016
 Phone (626) 386-1165 Fax (626) 386-1122
 Invoices to: Eurofins Eaton Analytical, LLC
 Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the Specified State Certification # and Exp Date for requested tests + matrix.
 Samples from: HAWAII

2-3 day rush

| Sample ID | Sample Date & Time Matrix | Clip Code | PWSID |
|--------------|---------------------------|-----------|-------|
| 202202161178 | 02/14/22 1026 DW | | JLS |

| Sample ID | Sample Date & Time Matrix | Clip Code | PWSID |
|--------------|---------------------------|-----------|-------|
| 202202161179 | 02/14/22 1026 DW | | JLS |

| | | |
|--|---------------------------|--|
| Type of Delivery <input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others | Airbill / Tracking Number | ECN <u>22 B180</u> |
| <input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery | | Recipient <u>Jocelyne Solis</u> |
| | | Date <u>02/17/22</u> Time <u>12:14</u> |

COC INSPECTION

| | | | | | |
|---|---|--|--|--|---|
| <input checked="" type="checkbox"/> Client Name | <input checked="" type="checkbox"/> Client PM/FC | <input type="checkbox"/> Sampler Name | <input checked="" type="checkbox"/> Sampling Date/Time | <input checked="" type="checkbox"/> Sample ID | <input type="checkbox"/> Matrix |
| <input checked="" type="checkbox"/> Address | <input checked="" type="checkbox"/> Tel # / Fax # | <input type="checkbox"/> Courier Signature | <input checked="" type="checkbox"/> Analysis Required | <input type="checkbox"/> Preservative (if any) | <input checked="" type="checkbox"/> TAT |
| Safety Issues (if any) | <input type="checkbox"/> High concentrations expected | <input type="checkbox"/> From Superfund Site | <input type="checkbox"/> Rad screening required | | |

Note: _____

PACKAGING INSPECTION

| | | | |
|---|--|--|--|
| Container | <input checked="" type="checkbox"/> Cooler | <input type="checkbox"/> Box | <input type="checkbox"/> Other |
| Condition | <input type="checkbox"/> Custody Seal | <input type="checkbox"/> Intact | <input type="checkbox"/> Damaged |
| Packaging | <input checked="" type="checkbox"/> Bubble Pack | <input type="checkbox"/> Styrofoam | <input type="checkbox"/> Popcorn |
| Temperatures (Cool, ≤6 °C but not frozen) | <input checked="" type="checkbox"/> Cooler 1 <u>2.4/1.9 °C</u> | <input checked="" type="checkbox"/> Cooler 2 <u>1.9/1.4 °C</u> | <input type="checkbox"/> Cooler 3 _____ °C |
| | <input type="checkbox"/> Cooler 6 _____ °C | <input type="checkbox"/> Cooler 7 _____ °C | <input checked="" type="checkbox"/> Cooler 4 <u>2.6/2.1 °C</u> |
| | <input type="checkbox"/> Cooler 8 _____ °C | <input type="checkbox"/> Cooler 9 _____ °C | <input type="checkbox"/> Cooler 5 _____ °C |
| | <input type="checkbox"/> Cooler 10 _____ °C | | <input type="checkbox"/> Cooler 10 _____ °C |

Thermometer: A - S/N 210191066 a 11/14 B - S/N 210271396 C - S/N 210271399 D - S/N _____

Comments: Temperature is out of range. PM was informed IMMEDIATELY.

Note: _____

DISCREPANCIES

| LabSampleID | LabSampleContainerID | Code | ClientSample Label ID / Information | Corrective Action |
|-------------|----------------------|------------|-------------------------------------|-------------------|
| <u>1</u> | <u>4-9</u> | <u>D22</u> | | <u>R8</u> |
| | | | | |

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time.

NOTES/OBSERVATIONS:

LEGEND:

| | | |
|--|---|---|
| <p>Code Description- Sample Management</p> <p>D1 Analysis is not indicated in _____</p> <p>D2 Analysis mismatch COC vs label</p> <p>D3 Sample ID mismatch COC vs label</p> <p>D4 Sample ID is not indicated in _____</p> <p>D5 Container -[improper] [leaking] [broken]</p> <p>D6 Date/Time is not indicated in _____</p> <p>D7 Date/Time mismatch COC vs label</p> <p>D8 Sample listed in COC is not received</p> <p>D9 Sample received is not listed in COC</p> <p>D10 No initial/date on corrections in COC/label</p> <p>D11 Container count mismatch COC vs received</p> <p>D12 Container size mismatch COC vs received</p> | <p>Code Description-Sample Management</p> <p>D13 Out of Holding Time</p> <p>D14 Bubble is >6mm</p> <p>D15 No trip blank in cooler</p> <p>D16 Preservation not indicated in _____</p> <p>D17 Preservation mismatch COC vs label</p> <p>D18 Insufficient chemical preservative</p> <p>D19 Insufficient Sample</p> <p>D20 No filtration info for dissolved analysis</p> <p>D21 No sample for moisture determination</p> <p><u>D22 Jct Fuel & Analysis not indicated on label</u></p> <p>D23 _____</p> <p>D24 _____</p> | <p><input type="checkbox"/> Continue to next page.</p> <p>Code Description-Sample Management</p> <p>R1 Proceed as indicated in <input type="checkbox"/> COC <input type="checkbox"/> Label</p> <p>R2 Refer to attached instruction</p> <p>R3 Cancel the analysis</p> <p>R4 Use vial with smallest bubble first</p> <p>R5 Log-in with latest sampling date and time+1 min</p> <p>R6 Adjust pH as necessary</p> <p>R7 Filter and preserved as necessary</p> <p>R8 <u>Informed Client.</u></p> <p>R9 _____</p> <p>R10 _____</p> <p>R11 _____</p> <p>R12 _____</p> |
|--|---|---|

REVIEWS:

Sample Labeling Maria Rivera SRF [Signature] PM [Signature]

Date 02/17/22 Date 2/17/22 Date 2/18/22

REPORTING CONVENTIONS

DATA QUALIFIERS:

| Lab Qualifier | AFCEE Qualifier | Description |
|---------------|-----------------|--|
| J | F | Indicates that the analyte is positively identified and the result is less than RL but greater than MDL. |
| N | | Indicates presumptive evidence of a compound. |
| B | B | Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level. |
| E | J | Indicates that the result is above the maximum calibration range or estimated value. |
| * | * | Out of QC limit. |

Note: The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

ACRONYMS AND ABBREVIATIONS:

| | |
|------|-----------------------------------|
| CRDL | Contract Required Detection Limit |
| RL | Reporting Limit |
| MRL | Method Reporting Limit |
| PQL | Practical Quantitation Limit |
| MDL | Method Detection Limit |
| DO | Diluted out |

DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

987974

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22B180

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 987974

SDG : 22B180

METHOD 5030B/8015B TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

A total of two(2) water samples were received on 02/17/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VG39B10B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VG39B10L/VG39B10C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Gasoline was within MS QC limits in B177-01M/B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

SAMPLE RESULTS

QC SUMMARIES

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987974
BATCH NO. : 22B180
METHOD : 5030B/8015B

| | | | |
|------------------|------------------|----------------|----------------|
| MATRIX | : WATER | | % MOISTURE:NA |
| DILUTION FACTOR: | 1 | 1 | 1 |
| SAMPLE ID | : MBLK1W | LCS1W | LCD1W |
| LAB SAMPLE ID | : VG39B10B | VG39B10L | VG39B10C |
| LAB FILE ID | : EB17005A | EB17006A | EB17007A |
| DATE PREPARED | : 02/17/22 16:15 | 02/17/22 16:52 | 02/17/22 17:28 |
| DATE ANALYZED | : 02/17/22 16:15 | 02/17/22 16:52 | 02/17/22 17:28 |
| PREP BATCH | : 22VG39B10 | 22VG39B10 | 22VG39B10 |
| CALIBRATION REF: | EB17003A | EB17003A | EB17003A |

ACCESSION:

| PARAMETERS | MBResult (mg/L) | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | SpikeAmt (mg/L) | LCDResult (mg/L) | LCDRec (%) | RPD (%) | QCLimit (%) | MaxRPD (%) |
|------------|--------------------|--------------------|---------------------|---------------|--------------------|---------------------|---------------|------------|----------------|---------------|
| Gasoline | ND | 0.500 | 0.443 | 89 | 0.500 | 0.456 | 91 | 3 | 60-130 | 30 |

| SURROGATE PARAMETER | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | SpikeAmt (mg/L) | LCDResult (mg/L) | LCDRec (%) | QCLimit (%) |
|---------------------|--------------------|---------------------|---------------|--------------------|---------------------|---------------|----------------|
| Bromofluorobenzene | 0.0400 | 0.0404 | 101 | 0.0400 | 0.0412 | 103 | 70-130 |

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 5030B/8015B

| | | | |
|------------------|------------------|----------------|-----------------|
| MATRIX | : WATER | | % MOISTURE:NA |
| DILUTION FACTOR: | 1 | 1 | 1 |
| SAMPLE ID | : 202202160931 | 202202160931MS | 202202160931MSD |
| LAB SAMPLE ID | : B177-01 | B177-01M | B177-01S |
| LAB FILE ID | : EB17011A | EB17012A | EB17013A |
| DATE PREPARED | : 02/17/22 19:53 | 02/17/22 20:30 | 02/17/22 21:06 |
| DATE ANALYZED | : 02/17/22 19:53 | 02/17/22 20:30 | 02/17/22 21:06 |
| PREP BATCH | : 22VG39B10 | 22VG39B10 | 22VG39B10 |
| CALIBRATION REF: | EB17003A | EB17003A | EB17003A |

ACCESSION:

| PARAMETERS | PSResult (mg/L) | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | RPD (%) | QCLimit (%) | MaxRPD (%) |
|------------|--------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|------------|----------------|---------------|
| Gasoline | ND | 0.500 | 0.495 | 99 | 0.500 | 0.499 | 100 | 1 | 50-130 | 30 |

| SURROGATE PARAMETER | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | QCLimit (%) |
|---------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|----------------|
| Bromofluorobenzene | 0.0400 | 0.0420 | 105 | 0.0400 | 0.0431 | 108 | 60-140 |

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

987974

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22B180

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 987974

SDG : 22B180

METHOD 3520C/8015B TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 02/17/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for Diesel was within LCS QC limits in DSB027WL. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. One(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 987974

SDG : 22B180

METHOD 3520C/8015B PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 02/17/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for JP5 was within LCS QC limits in J5B027WL. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. One(1) set of MS/MSD was analyzed. JP5 was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 987974

SDG : 22B180

METHOD 3520C/8015B PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 02/17/22 to be analyzed for Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSB027WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for JP8 was within LCS QC limits in J8B027WL. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. One(1) set of MS/MSD was analyzed. JP8 was within MS QC limits in 22B177-01M/22B177-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

Client : EUROFINS EATON ANALYTICAL
Project : 987974

SDG NO. : 22B180
Instrument ID : D5

| Client Sample ID | Laboratory Sample ID | Dilution Factor | % Moist | Analysis DateTime | Extraction DateTime | Sample Data FN | Calibration Data FN | Prep. Batch | Notes |
|------------------|----------------------|-----------------|---------|-------------------|---------------------|----------------|---------------------|-------------|--------------------------|
| MBLK1W | DSB027WB | 1 | NA | 02/22/2221:14 | 02/21/2210:30 | LB22012A | LB22006A | 22DSB027W | Method Blank |
| LCS1W | DSB027WL | 1 | NA | 02/22/2221:32 | 02/21/2210:30 | LB22013A | LB22006A | 22DSB027W | Lab Control Sample (LCS) |
| 202202161178 | B180-01 | 1 | NA | 02/23/2201:33 | 02/21/2210:30 | LB22026A | LB22006A | 22DSB027W | Field Sample |

FN - Filename
% Moist - Percent Moisture

SAMPLE RESULTS

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/14/22 10:26
Project     : 987974                     Date Received: 02/17/22
Batch No.   : 22B180                     Date Extracted: 02/21/22 10:30
Sample ID   : 202202161178              Date Analyzed: 02/23/22 01:33
Lab Samp ID : 22B180-01                  Dilution Factor: 1
Lab File ID : LB22026A                   Matrix: WATER
Ext Btch ID : 22DSB027W                  % Moisture: NA
Calib. Ref.: LB22006A                    Instrument ID: D5
=====

```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) |
|------------|-------------------|--------------|---------------|
| Diesel | ND | 0.026 | 0.013 |
| Motor Oil | ND | 0.052 | 0.026 |

| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT |
|----------------------|--------|---------|-----------|----------|
| Bromobenzene | 0.481 | 0.515 | 93 | 60-130 |
| Hexacosane | 0.123 | 0.129 | 96 | 60-130 |

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 970ml Final Volume : 5ml
Prepared by : POrreto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/14/22 10:26
Project     : 987974                     Date Received: 02/17/22
Batch No.   : 22B180                     Date Extracted: 02/21/22 10:30
Sample ID   : 202202161178              Date Analyzed: 02/23/22 01:33
Lab Samp ID: 22B180-01                   Dilution Factor: 1
Lab File ID: LB22026A                    Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22007A                    Instrument ID: D5
=====
    
```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) | |
|----------------------|-------------------|--------------|---------------|----------|
| JP5 | ND | 0.052 | 0.026 | |
| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT |
| Bromobenzene | 0.481 | 0.515 | 93 | 60-130 |
| Hexacosane | 0.123 | 0.129 | 96 | 60-130 |

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP5 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 970ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/14/22 10:26
Project    : 987974                       Date Received: 02/17/22
Batch No.  : 22B180                       Date Extracted: 02/21/22 10:30
Sample ID  : 202202161178                 Date Analyzed: 02/23/22 01:33
Lab Samp ID: 22B180-01                   Dilution Factor: 1
Lab File ID: LB22026A                     Matrix: WATER
Ext Btch ID: 22DSB027W                    % Moisture: NA
Calib. Ref.: LB22008A                     Instrument ID: D5
=====
  
```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) | |
|----------------------|-------------------|--------------|---------------|----------|
| JP8 | ND | 0.052 | 0.026 | |
| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT |
| Bromobenzene | 0.481 | 0.515 | 93 | 60-130 |
| Hexacosane | 0.123 | 0.129 | 96 | 60-130 |

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP8 C8-C18
 Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
 Sample Amount : 970ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

QC SUMMARIES

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/21/22 10:30
Project    : 987974                       Date Received: 02/21/22
Batch No.  : 22B180                       Date Extracted: 02/21/22 10:30
Sample ID  : MBLK1W                       Date Analyzed: 02/22/22 21:14
Lab Samp ID: DSB027WB                     Dilution Factor: 1
Lab File ID: LB22012A                     Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22006A                    Instrument ID: D5
=====

```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) |
|------------|-------------------|--------------|---------------|
| Diesel | ND | 0.025 | 0.012 |
| Motor Oil | ND | 0.050 | 0.025 |

| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT |
|----------------------|--------|---------|-----------|----------|
| Bromobenzene | 0.384 | 0.500 | 77 | 60-130 |
| Hexacosane | 0.110 | 0.125 | 88 | 60-130 |

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
Prepared by : P0reto Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987974
BATCH NO. : 22B180
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : MBLK1W LCS1W
LAB SAMPLE ID : DSB027WB DSB027WL
LAB FILE ID : LB22012A LB22013A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 21:14 02/22/22 21:32
PREP BATCH : 22DSB027W 22DSB027W
CALIBRATION REF: LB22006A LB22006A

ACCESSION:

| PARAMETERS | MBResult (mg/L) | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|------------|--------------------|--------------------|---------------------|---------------|----------------|
| Diesel | ND | 2.50 | 2.72 | 109 | 50-130 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|----------------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.500 | 0.567 | 113 | 60-130 |
| Hexacosane | 0.125 | 0.130 | 104 | 60-130 |

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 202202160931                       202202160931MSD
LAB SAMPLE ID : 22B177-01                         22B177-01S
LAB FILE ID  : LB22017A                          LB22019A
DATE PREPARED : 02/21/22 10:30                  02/21/22 10:30
DATE ANALYZED : 02/22/22 22:46                  02/22/22 23:23
PREP BATCH   : 22DSB027W                        22DSB027W
CALIBRATION REF: LB22006A                       LB22006A
  
```

ACCESSION:

| PARAMETERS | PSResult (mg/L) | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | RPD (%) | QCLimit (%) | MaxRPD (%) |
|------------|--------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|------------|----------------|---------------|
| Diesel | ND | 2.50 | 2.89 | 116 | 2.55 | 3.16 | 124 | 9 | 50-130 | 30 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | QCLimit (%) |
|----------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.500 | 0.478 | 96 | 0.510 | 0.512 | 100 | 60-130 |
| Hexacosane | 0.125 | 0.122 | 98 | 0.127 | 0.132 | 104 | 60-130 |

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/21/22 10:30
Project     : 987974                      Date Received: 02/21/22
Batch No.   : 22B180                      Date Extracted: 02/21/22 10:30
Sample ID   : MBLK1W                     Date Analyzed: 02/22/22 21:14
Lab Samp ID : DSB027WB                   Dilution Factor: 1
Lab File ID : LB22012A                   Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22007A                     Instrument ID: D5
=====
  
```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) | | |
|----------------------|-------------------|--------------|---------------|----------|--|
| JP5 | ND | 0.050 | 0.025 | | |
| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT | |
| Bromobenzene | 0.384 | 0.500 | 77 | 60-130 | |
| Hexacosane | 0.110 | 0.125 | 88 | 60-130 | |

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP5 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
 Prepared by : POrreto Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987974
BATCH NO. : 22B180
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : MBLK1W LCS1W
LAB SAMPLE ID : DSB027WB J5B027WL
LAB FILE ID : LB22012A LB22014A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 21:14 02/22/22 21:51
PREP BATCH : 22DSB027W 22DSB027W
CALIBRATION REF: LB22007A LB22007A

ACCESSION:

| PARAMETERS | MBResult (mg/L) | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|------------|--------------------|--------------------|---------------------|---------------|----------------|
| JP5 | ND | 2.50 | 2.25 | 90 | 30-160 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|----------------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.500 | 0.450 | 90 | 60-130 |
| Hexacosane | 0.125 | 0.119 | 95 | 60-130 |

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

| | | | |
|------------------|------------------|----------------|-----------------|
| MATRIX | : WATER | | % MOISTURE:NA |
| DILUTION FACTOR: | 1 | 1 | 1 |
| SAMPLE ID | : 202202160931 | 202202160931MS | 202202160931MSD |
| LAB SAMPLE ID | : 22B177-01 | 22B177-01M | 22B177-01S |
| LAB FILE ID | : LB22017A | LB22020A | LB22021A |
| DATE PREPARED | : 02/21/22 10:30 | 02/21/22 10:30 | 02/21/22 10:30 |
| DATE ANALYZED | : 02/22/22 22:46 | 02/22/22 23:42 | 02/23/22 00:00 |
| PREP BATCH | : 22DSB027W | 22DSB027W | 22DSB027W |
| CALIBRATION REF: | LB22007A | LB22007A | LB22007A |

ACCESSION:

| PARAMETERS | PSResult (mg/L) | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | RPD (%) | QCLimit (%) | MaxRPD (%) |
|------------|--------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|------------|----------------|---------------|
| JP5 | ND | 2.53 | 2.69 | 107 | 2.53 | 2.59 | 103 | 4 | 30-160 | 30 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | QCLimit (%) |
|----------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.505 | 0.495 | 98 | 0.505 | 0.501 | 99 | 60-130 |
| Hexacosane | 0.126 | 0.121 | 96 | 0.126 | 0.120 | 95 | 60-130 |

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

METHOD 3520C/8015B
 PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 02/21/22 10:30
Project     : 987974                     Date Received: 02/21/22
Batch No.   : 22B180                     Date Extracted: 02/21/22 10:30
Sample ID   : MBLK1W                     Date Analyzed: 02/22/22 21:14
Lab Samp ID: DSB027WB                    Dilution Factor: 1
Lab File ID: LB22012A                    Matrix: WATER
Ext Btch ID: 22DSB027W                   % Moisture: NA
Calib. Ref.: LB22008A                    Instrument ID: D5
=====
  
```

| PARAMETERS | RESULTS (mg/L) | RL (mg/L) | MDL (mg/L) |
|------------|-------------------|--------------|---------------|
| JP8 | ND | 0.050 | 0.025 |

| SURROGATE PARAMETERS | RESULT | SPK_AMT | %RECOVERY | QC LIMIT |
|----------------------|--------|---------|-----------|----------|
| Bromobenzene | 0.384 | 0.500 | 77 | 60-130 |
| Hexacosane | 0.110 | 0.125 | 88 | 60-130 |

Notes:

RL : Reporting Limit
 Parameter H-C Range
 JP8 C8-C18

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
 Prepared by : POrto Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987974
BATCH NO. : 22B180
METHOD : 3520C/8015B

MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1
SAMPLE ID : MBLK1W LCS1W
LAB SAMPLE ID : DSB027WB J8B027WL
LAB FILE ID : LB22012A LB22015A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 21:14 02/22/22 22:09
PREP BATCH : 22DSB027W 22DSB027W
CALIBRATION REF: LB22008A LB22008A

ACCESSION:

| PARAMETERS | MBResult (mg/L) | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|------------|--------------------|--------------------|---------------------|---------------|----------------|
| JP8 | ND | 2.50 | 2.21 | 88 | 30-160 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | LCSResult (mg/L) | LCSRec (%) | QCLimit (%) |
|----------------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.500 | 0.528 | 106 | 60-130 |
| Hexacosane | 0.125 | 0.122 | 98 | 60-130 |

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 987883
BATCH NO. : 22B177
METHOD : 3520C/8015B

```

=====
MATRIX : WATER % MOISTURE:NA
DILUTION FACTOR: 1 1 1
SAMPLE ID : 202202160931 202202160931MS 202202160931MSD
LAB SAMPLE ID : 22B177-01 22B177-01M 22B177-01S
LAB FILE ID : LB22017A LB22022A LB22023A
DATE PREPARED : 02/21/22 10:30 02/21/22 10:30 02/21/22 10:30
DATE ANALYZED : 02/22/22 22:46 02/23/22 00:19 02/23/22 00:38
PREP BATCH : 22DSB027W 22DSB027W 22DSB027W
CALIBRATION REF: LB22008A LB22008A LB22008A
  
```

ACCESSION:

| PARAMETERS | PSResult (mg/L) | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | RPD (%) | QCLimit (%) | MaxRPD (%) |
|------------|--------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|------------|----------------|---------------|
| JP8 | ND | 2.62 | 2.58 | 98 | 2.62 | 2.26 | 86 | 13 | 30-160 | 30 |

| SURROGATE PARAMETERS | SpikeAmt (mg/L) | MSResult (mg/L) | MSRec (%) | SpikeAmt (mg/L) | MSDResult (mg/L) | MSDRec (%) | QCLimit (%) |
|----------------------|--------------------|--------------------|--------------|--------------------|---------------------|---------------|----------------|
| Bromobenzene | 0.525 | 0.563 | 107 | 0.525 | 0.491 | 94 | 60-130 |
| Hexacosane | 0.131 | 0.128 | 98 | 0.131 | 0.123 | 94 | 60-130 |

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate